



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102013

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/10/2019	17:00
2	CAL2	1	Water	NA	NA	20.2	7.00	04/10/2019	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	04/10/2019	17:02
4	ICV	1	Water	NA	NA	20.2	6.99	04/10/2019	17:05
5	CCV1	1	Water	NA	NA	20.2	2.02	04/10/2019	17:06
6	K2355-04	1	Solid	20.03	20	22.4	8.29	04/10/2019	17:08
7	K2355-04DUP	1	Solid	20.04	20	22.5	8.30	04/10/2019	17:10
8	K2358-04	1	Solid	20.02	20	22.5	8.23	04/10/2019	17:11
9	K2358-08	1	Solid	20.01	20	22.7	7.39	04/10/2019	17:12
10	K2358-12	1	Solid	20.03	20	22.4	7.71	04/10/2019	17:15
11	K2358-16	1	Solid	20.02	20	22.5	9.54	04/10/2019	17:16
12	K2359-01	1	Solid	20.03	20	21.2	5.56	04/10/2019	17:19
13	CCV2	1	Water	NA	NA	20.3	12.01	04/10/2019	17:20

LB102013

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2347

WorkList ID : 124656

Date : 04-10-2019 15:35:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2019	Solid	K2355-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-9-S	04/10/2019	9045D
04/10/2019	Solid	K2358-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-5	04/10/2019	9045D
04/09/2019	Solid	K2358-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-9	04/09/2019	9045D
04/09/2019	Solid	K2358-12	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-7-S	04/09/2019	9045D
04/09/2019	Solid	K2358-16	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-7-D	04/09/2019	9045D
04/08/2019	Solid	K2359-01	Corrosivity	Cool 4 deg C	CLEA11	K11	OIL-WATER-SEPERATOR-2	04/08/2019	9045D

Date/Time 04-10-19 15:40
 Received by: JP
 Relinquished by: CA

Date/Time 04-10-19 18:00
 Received by: JP
 Relinquished by: JP