



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102018

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	04/10/2019	08:45
2	K2335-02		1	9	>100.0	>212.0	>212.0	04/10/2019	09:20
3	K2335-02DUP		1	9	>100.0	>212.0	>212.0	04/10/2019	09:40
4	K2342-18		1	7	>100.0	>212.0	>212.0	04/10/2019	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102018

WorkList Name : FP41019

WorkList ID : 124624

Date : 04-10-2019 08:19:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2019	Water	K2335-02	C Flash Point	Cool 4 deg C	PSEG01	K31	SV30755L	04/09/2019	1010A
04/15/2019	Water	K2342-18	C Flash Point	Cool 4 deg C	PSEG01	K31	TP-12	04/08/2019	1010A

Date/Time 04/10/19 08:20
 Received by: KS
 Relinquished by: CP

Date/Time 04/10/19 10:15
 Received by: CP
 Relinquished by: KS