

LB102022


245.1

INSTRUMENT ID: CU1

Linear

 μ Abs.:

40996



A= 0.0000e+000

B= 2.4587e-004

C= -1.2412e-002

Rho= 0.9990254

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	57	0.000	57				
0.05	0.050	0.039	-0.011	211	0.0 %	211				
0.2	0.200	0.180	-0.020	781	0.0 %	781				
2.5	2.500	2.464	-0.036	10072	0.0 %	10072				
5.0	5.000	5.300	0.300	21609	0.0 %	21609				
7.5	7.500	7.198	-0.302	29326	0.0 %	29326				
10.0	10.000	10.067	0.067	40996	0.0 %	40996				

LB102022 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Apr 2019 09:33:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	57	-	0.0000	245.1	PPB	11 Apr 2019 09:36:08	S	Std
0.05 - 1	50-05	211	-	0.0500	245.1	PPB	11 Apr 2019 09:38:25	S	Std
0.2 - 1	50-2	781	-	0.2000	245.1	PPB	11 Apr 2019 09:40:43	S	Std
2.5 - 1	52-5	10072	-	2.5000	245.1	PPB	11 Apr 2019 09:43:00	S	Std
5.0 - 1	55-0	21609	-	5.0000	245.1	PPB	11 Apr 2019 09:45:17	S	Std
7.5 - 1	57-5	29326	-	7.5000	245.1	PPB	11 Apr 2019 09:47:33	S	Std
10.0 - 1	510-0	40996	-	10.0000	245.1	PPB	11 Apr 2019 09:49:50	S	Std
ICV94 - 1	ICV94	16594	4.0675	-245.1		PPB	11 Apr 2019 10:01:16	U	SMPL
ICB94 - 1	ICB94	-17	-0.0166	-245.1		PPB	11 Apr 2019 10:03:32	U	SMPL
CCV77 - 1	CCV77	21068	5.1675	-245.1		PPB	11 Apr 2019 10:05:47	U	SMPL
CCB77 - 1	CCB77	-44	-0.0232	-245.1		PPB	11 Apr 2019 10:08:03	U	SMPL
CRA - 1	CRA	198	0.0363	-245.1		PPB	11 Apr 2019 10:10:18	U	SMPL
HighStd - 1	HighStd	41359	10.1563	-245.1		PPB	11 Apr 2019 10:12:33	U	SMPL
ChkStd - 1	ChkStd	28591	7.0171	-245.1		PPB	11 Apr 2019 10:14:49	U	SMPL
PB118741BL - 1	PBW	-74	-0.0306	-245.1		PPB	11 Apr 2019 10:17:07	U	SMPL
PB118741BS - 1	LCSW	16519	4.0490	-245.1		PPB	11 Apr 2019 10:19:23	U	SMPL
K2279-01 - 1	32220008-3	169	0.0291	-245.1		PPB	11 Apr 2019 10:21:40	U	SMPL
K2279-01DUP - 1	32220008-3DUP	224	0.0427	-245.1		PPB	11 Apr 2019 10:23:56	U	SMPL
K2279-01MS - 1	32220008-3MS	16827	4.1248	-245.1		PPB	11 Apr 2019 10:26:13	U	SMPL
K2279-01MSD - 1	32220008-3MSD	16663	4.0844	-245.1		PPB	11 Apr 2019 10:28:31	U	SMPL
K2331-01 - 1	T1-604-EB	224	0.0427	-245.1		PPB	11 Apr 2019 10:30:45	U	SMPL
CCV78 - 1	CCV78	21168	5.1921	-245.1		PPB	11 Apr 2019 10:33:02	U	SMPL
CCB78 - 1	CCB78	-20	-0.0173	-245.1		PPB	11 Apr 2019 10:35:17	U	SMPL
K2331-03 - 1	T1-604	261	0.0518	-245.1		PPB	11 Apr 2019 10:37:33	U	SMPL
K2331-04 - 1	T1-604-X	326	0.0677	-245.1		PPB	11 Apr 2019 10:39:48	U	SMPL
K2332-02 - 1	OH-607	266	0.0530	-245.1		PPB	11 Apr 2019 10:42:04	U	SMPL
K2333-03 - 1	NCQ-632	276	0.0554	-245.1		PPB	11 Apr 2019 10:44:19	U	SMPL
K2279-01LX5 - 1		87	0.0090	-245.1		PPB	11 Apr 2019 10:46:35	U	SMPL
K2279-01A - 1		16680	4.0886	-245.1		PPB	11 Apr 2019 10:48:52	U	SMPL
CCV79 - 1	CCV79	20668	5.0691	-245.1		PPB	11 Apr 2019 10:51:08	U	SMPL
CCB79 - 1	CCB79	-17	-0.0166	-245.1		PPB	11 Apr 2019 10:53:26	U	SMPL