

LB102040

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/11/2019 15:13

Test: Total CN

ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.541	0.0	0.088	
ICB1	0.222	0.0	0.004	
CCV1	240.319	0.0	0.218	
CCB1	0.066	0.0	0.004	
PB118801BL	-0.291	0.0	0.004	
PB118801BS	98.184	0.0	0.092	
MIDPB118801	247.227	0.0	0.224	
K2241-12	1.315	0.0	0.005	
CCV2	249.147	0.0	0.226	
CCB2	0.191	0.0	0.004	

N	10
Mean	92.992
SD	111.9847
CV%	120.42

Aquakem v. 7.2AQ1

Results from time period:

Thu Apr 11 15:12:55 2019

Thu Apr 11 15:13:04 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.0664	µg/l	4/11/2019 12:32:29	
5.OPPBCN	A	Total CN	P	4.405	µg/l	4/11/2019 12:32:30	
10PPBCN	A	Total CN	P	9.8979	µg/l	4/11/2019 12:32:31	
50PPBCN	A	Total CN	P	50.2383	µg/l	4/11/2019 12:32:32	
100PPBCN	A	Total CN	P	100.6053	µg/l	4/11/2019 12:32:33	
250PPBCN	A	Total CN	P	250.1136	µg/l	4/11/2019 12:32:34	
500PPBCN	A	Total CN	P	499.8063	µg/l	4/11/2019 12:32:35	
ICV1	S	Total CN	P	93.5406	µg/l	4/11/2019 15:12:55	
ICB1	S	Total CN	P	0.2223	µg/l	4/11/2019 15:12:56	
CCV1	S	Total CN	P	240.3193	µg/l	4/11/2019 15:12:57	
CCB1	S	Total CN	P	0.0663	µg/l	4/11/2019 15:12:58	
PB118801BL	S	Total CN	P	-0.2909	µg/l	4/11/2019 15:12:59	
PB118801BS	S	Total CN	P	98.1837	µg/l	4/11/2019 15:13:00	
MIDPB118801	S	Total CN	P	247.2269	µg/l	4/11/2019 15:13:01	
K2241-12	S	Total CN	P	1.3154	µg/l	4/11/2019 15:13:02	
CCV2	S	Total CN	P	249.1472	µg/l	4/11/2019 15:13:03	
CCB2	S	Total CN	P	0.1911	µg/l	4/11/2019 15:13:04	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

Instrument ID : Konelab

4/11/2019 12:36


Test Total CN

Accepted 4/11/2019 12:36

Factor 1123
Bias 0.004

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0664	0.0000	
2	5.0PPBCN	0.008	4.4050	5.0000	
3	10PPBCN	0.013	9.8979	10.0000	
4	50PPBCN	0.049	50.2383	50.0000	
5	100PPBCN	0.094	100.6053	100.0000	
6	250PPBCN	0.227	250.1136	250.0000	
7	500PPBCN	0.449	499.8063	500.0000	