



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102042

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/11/2019	15:35
2	CAL2	1	Water	NA	NA	20.1	7.01	04/11/2019	15:36
3	CAL3	1	Water	NA	NA	20.1	10.02	04/11/2019	15:37
4	ICV	1	Water	NA	NA	20.1	7.00	04/11/2019	15:38
5	CCV1	1	Water	NA	NA	20.1	2.01	04/11/2019	15:40
6	K2364-01	1	Water	NA	NA	13.4	5.71	04/11/2019	15:41
7	K2364-01DUP	1	Water	NA	NA	13.5	5.72	04/11/2019	15:42
8	K2364-02	1	Water	NA	NA	13.8	5.71	04/11/2019	15:43
9	K2371-01	1	Water	NA	NA	14.4	5.69	04/11/2019	15:46
10	K2381-01	1	Water	NA	NA	14.7	5.36	04/11/2019	15:48
11	K2381-02	1	Water	NA	NA	15.0	5.00	04/11/2019	15:49
12	K2381-03	1	Water	NA	NA	15.6	5.09	04/11/2019	15:50
13	CCV2	1	Water	NA	NA	20.2	12.01	04/11/2019	15:52

19/10/2042

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k2171

WorkList ID : 124720

Date : 04-11-2019 15:27:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2019	Water	K2364-01 -C	pH	Cool 4 deg C	PSEG01	K31	OR-COMP-1	04/11/2019	9040C
04/11/2019	Water	K2364-02 -C	pH	Cool 4 deg C	PSEG01	K31	OR-COMP-2	04/11/2019	9040C
04/11/2019	Water	K2371-01 -C	pH	Cool 4 deg C	PSEG01	K31	OK-COMP	04/11/2019	9040C
04/11/2019	Water	K2381-01 -C	pH	Cool 4 deg C	PSEG01	K11	CL-COMP-1	04/11/2019	9040C
04/11/2019	Water	K2381-02 -C	pH	Cool 4 deg C	PSEG01	K11	CL-COMP-2	04/11/2019	9040C
04/11/2019	Water	K2381-03 -C	pH	Cool 4 deg C	PSEG01	K11	CL-COMP-3	04/11/2019	9040C

Date/Time 04.11.19 15:30
 Received by: JD
 Relinquished by: EO

Date/Time 04.11.19 16:25
 Received by: EN
 Relinquished by: JD