



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102045

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/11/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/11/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/11/2019	17:17
4	ICV	1	Water	NA	NA	20.1	6.99	04/11/2019	17:18
5	CCV1	1	Water	NA	NA	20.1	2.01	04/11/2019	17:20
6	K2196-04	1	Solid	20.02	20	21.5	6.83	04/11/2019	17:21
7	K2204-02	1	Solid	20.03	20	21.4	8.63	04/11/2019	17:22
8	K2373-03	1	Solid	20.05	20	21.8	8.13	04/11/2019	17:25
9	K2374-03	1	Solid	20.04	20	21.3	7.82	04/11/2019	17:27
10	K2376-04	1	Solid	20.03	20	21.7	8.37	04/11/2019	17:28
11	K2377-04	1	Solid	20.02	20	21.6	7.37	04/11/2019	17:30
12	K2377-08	1	Solid	20.03	20	21.7	8.40	04/11/2019	17:31
13	K2377-08DUP	1	Solid	20.04	20	21.6	8.41	04/11/2019	17:33
14	CCV2	1	Water	NA	NA	20.1	12.01	04/11/2019	17:35

LB102045

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2196

WorkList ID : 124717

Date : 04-11-2019 15:24:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2019	Solid	K2196-04	Corrosivity	Cool 4 deg C	PSEG05		HD-02-041119	04/11/2019	9045D
04/11/2019	Solid	K2204-02	Corrosivity	Cool 4 deg C	PSEG05	K31	TR-04-041119	04/11/2019	9045D
04/11/2019	Solid	K2373-03	Corrosivity	Cool 4 deg C	PSEG01	K31	TPH-FOUND	04/11/2019	9045D
04/11/2019	Solid	K2374-03	Corrosivity	Cool 4 deg C	WSP001	K31	P-1(0-10)	04/11/2019	9045D
04/10/2019	Solid	K2376-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-4-S	04/10/2019	9045D
04/10/2019	Solid	K2377-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-13-S	04/10/2019	9045D
04/10/2019	Solid	K2377-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-12	04/10/2019	9045D

Date/Time 04.11.19 16:10
 Received by: Jh
 Relinquished by: CC

Date/Time 04.11.19 17:35
 Received by: CH
 Relinquished by: Jh