

LB102057


7470A

INSTRUMENT ID: CU1

Linear

 μ Abs.:

40613



A= 0.0000e+000

B= 2.4074e-004

C= -3.8288e-002

Rho= 0.9989971

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	28	0.000	28				
0.2	0.200	0.159	-0.041	818	0.0 %	818				
2.5	2.500	2.487	-0.013	10491	0.0 %	10491				
5.0	5.000	5.045	0.045	21116	0.0 %	21116				
7.5	7.500	7.801	0.301	32564	0.0 %	32564				
10.0	10.000	9.739	-0.261	40613	0.0 %	40613				

LB102057 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Apr 2019 13:56:21

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	28	-	0.0000	7470A	PPB	12 Apr 2019 14:03:45	S	Std
0.2 - 1	50-2	818	-	0.2000	7470A	PPB	12 Apr 2019 14:05:54	S	Std
2.5 - 1	52-5	10491	-	2.5000	7470A	PPB	12 Apr 2019 14:08:02	S	Std
5.0 - 1	55-0	21116	-	5.0000	7470A	PPB	12 Apr 2019 14:10:10	S	Std
7.5 - 1	57-5	32564	-	7.5000	7470A	PPB	12 Apr 2019 14:12:17	S	Std
10.0 - 1	510-0	40613	-	10.0000	7470A	PPB	12 Apr 2019 14:18:18	S	Std
ICV98 - 1	ICV98	16528	3.9407	-7470A	PPB	12 Apr 2019 14:23:05	U	SMPL	
ICB98 - 1	ICB98	-7	-0.0400	-7470A	PPB	12 Apr 2019 14:25:11	U	SMPL	
CCV94 - 1	CCV94	20971	5.0103	-7470A	PPB	12 Apr 2019 14:27:19	U	SMPL	
CCB94 - 1	CCB94	-19	-0.0429	-7470A	PPB	12 Apr 2019 14:29:25	U	SMPL	
CRA - 1	CRA	955	0.1916	-7470A	PPB	12 Apr 2019 14:35:55	U	SMPL	
HighStd - 1	HighStd	40626	9.7422	-7470A	PPB	12 Apr 2019 14:38:01	U	SMPL	
ChkStd - 1	ChkStd	29975	7.1780	-7470A	PPB	12 Apr 2019 14:40:08	U	SMPL	
PB118853BL - 1	PBW	-53	-0.0510	-7470A	PPB	12 Apr 2019 14:42:15	U	SMPL	
PB118853BS - 1	LCSW	17655	4.2120	-7470A	PPB	12 Apr 2019 14:44:22	U	SMPL	
K2355-04 - 1	TP-9-S	245	0.0207	-7470A	PPB	12 Apr 2019 14:46:31	U	SMPL	
K2355-04DUP - 1	TP-9-SDUP	272	0.0272	-7470A	PPB	12 Apr 2019 14:48:38	U	SMPL	
K2355-04MS - 1	TP-9-SMS	17638	4.2079	-7470A	PPB	12 Apr 2019 14:50:46	U	SMPL	
K2355-04MSD - 1	TP-9-SMSD	17484	4.1709	-7470A	PPB	12 Apr 2019 14:52:54	U	SMPL	
K2357-01 - 1	CS-041019	83	-0.0183	-7470A	PPB	12 Apr 2019 14:55:00	U	SMPL	
CCV95 - 1	CCV95	20817	4.9733	-7470A	PPB	12 Apr 2019 14:57:06	U	SMPL	
CCB95 - 1	CCB95	-18	-0.0426	-7470A	PPB	12 Apr 2019 14:59:12	U	SMPL	
K2358-04 - 1	TP-5	72	-0.0210	-7470A	PPB	12 Apr 2019 15:01:19	U	SMPL	
K2358-08 - 1	TP-9	223	0.0154	-7470A	PPB	12 Apr 2019 15:03:25	U	SMPL	
K2358-12 - 1	TP-7-S	331	0.0414	-7470A	PPB	12 Apr 2019 15:05:32	U	SMPL	
K2358-16 - 1	TP-7-D	334	0.0421	-7470A	PPB	12 Apr 2019 15:07:38	U	SMPL	
K2359-01 - 1	OIL-WATER-SEPERATOR-2	182	0.0055	-7470A	PPB	12 Apr 2019 15:09:46	U	SMPL	
PB118859BL - 1	PBW	17	-0.0342	-7470A	PPB	12 Apr 2019 15:11:53	U	SMPL	
PB118859BS - 1	LCSW	17641	4.2087	-7470A	PPB	12 Apr 2019 15:14:01	U	SMPL	
K2196-04 - 1	HD-02-041119	-11	-0.0409	-7470A	PPB	12 Apr 2019 15:16:11	U	SMPL	
K2196-04DUP - 1	HD-02-041119DUP	27	-0.0318	-7470A	PPB	12 Apr 2019 15:18:19	U	SMPL	
K2196-04MS - 1	HD-02-041119MS	17838	4.2561	-7470A	PPB	12 Apr 2019 15:20:27	U	SMPL	
CCV96 - 1	CCV96	21429	5.1206	-7470A	PPB	12 Apr 2019 15:22:44	U	SMPL	
CCB96 - 1	CCB96	-9	-0.0405	-7470A	PPB	12 Apr 2019 15:24:58	U	SMPL	
K2196-04MSD - 1	HD-02-041119MSD	17696	4.2219	-7470A	PPB	12 Apr 2019 15:27:16	U	SMPL	
K2204-02 - 1	TR-04-041119	-26	-0.0445	-7470A	PPB	12 Apr 2019 15:29:25	U	SMPL	
K2373-03 - 1	TPH-FOUND	10	-0.0359	-7470A	PPB	12 Apr 2019 15:31:32	U	SMPL	
K2374-03 - 1	P-1(0-10)	53	-0.0255	-7470A	PPB	12 Apr 2019 15:33:39	U	SMPL	
K2376-04 - 1	TP-4-S	60	-0.0238	-7470A	PPB	12 Apr 2019 15:35:45	U	SMPL	
K2377-04 - 1	TP-13-S	1	-0.0380	-7470A	PPB	12 Apr 2019 15:37:53	U	SMPL	
K2377-08 - 1	TP-12-S	13	-0.0352	-7470A	PPB	12 Apr 2019 15:40:00	U	SMPL	
PB118860BL - 1	PBW	9	-0.0361	-7470A	PPB	12 Apr 2019 15:42:07	U	SMPL	
PB118860BS - 1	LCSW	17652	4.2113	-7470A	PPB	12 Apr 2019 15:44:16	U	SMPL	
K2364-01 - 1	OR-COMP-1	307	0.0356	-7470A	PPB	12 Apr 2019 15:46:23	U	SMPL	
CCV97 - 1	CCV97	22132	5.2898	-7470A	PPB	12 Apr 2019 15:48:29	U	SMPL	
CCB97 - 1	CCB97	-43	-0.0486	-7470A	PPB	12 Apr 2019 15:50:36	U	SMPL	
K2364-01DUP - 1	OR-COMP-1DUP	205	0.0111	-7470A	PPB	12 Apr 2019 15:52:42	U	SMPL	
K2364-01MS - 1	OR-COMP-1MS	16615	3.9617	-7470A	PPB	12 Apr 2019 15:54:48	U	SMPL	
K2364-01MSD - 1	OR-COMP-1MSD	16732	3.9898	-7470A	PPB	12 Apr 2019 15:56:55	U	SMPL	
K2364-02 - 1	OR-COMP-2	191	0.0077	-7470A	PPB	12 Apr 2019 15:59:02	U	SMPL	
K2371-01 - 1	OK-COMP	260	0.0243	-7470A	PPB	12 Apr 2019 16:01:09	U	SMPL	
K2381-01 - 1	CL-COMP-1	-18	-0.0426	-7470A	PPB	12 Apr 2019 16:03:16	U	SMPL	
K2381-02 - 1	CL-COMP-2	4890	1.1389	-7470A	PPB	12 Apr 2019 16:05:24	U	SMPL	
K2381-03 - 1	CL-COMP-3	5634	1.3181	-7470A	PPB	12 Apr 2019 16:07:32	U	SMPL	
PB118853TB - 1	PB118799TB	-49	-0.0501	-7470A	PPB	12 Apr 2019 16:09:40	U	SMPL	
PB118859TB - 1	PB118833TB	79	-0.0193	-7470A	PPB	12 Apr 2019 16:11:47	U	SMPL	
CCV98 - 1	CCV98	21932	5.2417	-7470A	PPB	12 Apr 2019 16:13:53	U	SMPL	
CCB98 - 1	CCB98	-17	-0.0424	-7470A	PPB	12 Apr 2019 16:16:02	U	SMPL	
K2355-04LX5 - 1		72	-0.0210	-7470A	PPB	12 Apr 2019 16:18:09	U	SMPL	
K2355-04A - 1		17410	4.1531	-7470A	PPB	12 Apr 2019 16:20:16	U	SMPL	
K2196-04LX5 - 1		-11	-0.0409	-7470A	PPB	12 Apr 2019 16:22:22	U	SMPL	
K2196-04A - 1		17465	4.1663	-7470A	PPB	12 Apr 2019 16:24:29	U	SMPL	
K2364-01LX5 - 1		67	-0.0222	-7470A	PPB	12 Apr 2019 16:26:36	U	SMPL	
K2364-01A - 1		16889	4.0276	-7470A	PPB	12 Apr 2019 16:28:43	U	SMPL	
CCV99 - 1	CCV99	22121	5.2872	-7470A	PPB	12 Apr 2019 16:30:50	U	SMPL	
CCB99 - 1	CCB99	-73	-0.0559	-7470A	PPB	12 Apr 2019 16:32:58	U	SMPL	