



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB102064

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/12/2019	16:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/12/2019	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/12/2019	16:18
4	ICV	1	Water	NA	NA	20.2	7.00	04/12/2019	16:20
5	CCV1	1	Water	NA	NA	20.1	2.01	04/12/2019	16:21
6	K2397-18	1	Water	NA	NA	13.6	6.73	04/12/2019	16:22
7	K2397-18DUP	1	Water	NA	NA	13.7	6.74	04/12/2019	16:24
8	CCV2	1	Water	NA	NA	20.2	12.02	04/12/2019	16:25

102064

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2397

WorkList ID : 124784

Date : 04-12-2019 16:07:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2019	Water	K2397-18-17	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-3	04/11/2019	9040C

Date/Time 04-12-19 16:10
Received by: In
Relinquished by: DP

Date/Time 04-12-19 17:00
Received by: DP
Relinquished by: JD