



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102072

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/12/2019	14:02
2	K2364-01		1	7	100.00	>212.0	>212.0	04/12/2019	14:35
3	K2364-01DUP		1	7	100.00	>212.0	>212.0	04/12/2019	15:05
4	K2364-02		1	9	100.00	>212.0	>212.0	04/12/2019	15:35
5	K2371-01		1	7	100.00	>212.0	>212.0	04/12/2019	15:55
6	K2381-01		1	9	100.00	>212.0	>212.0	04/12/2019	16:15
7	K2381-02		1	6	100.00	>212.0	>212.0	04/12/2019	16:40
8	K2381-03		1	7	100.00	>212.0	>212.0	04/12/2019	16:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102072

WorkList Name : FP41219

WorkList ID : 124773

Date : 04-12-2019 13:38:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/18/2019	Water	K2364-01	Flash Point	Cool 4 deg C	PSEG01	K31	OR-COMP-1	04/11/2019	1010A
04/18/2019	Water	K2364-02	Flash Point	Cool 4 deg C	PSEG01	K31	OR-COMP-2	04/11/2019	1010A
04/18/2019	Water	K2371-01	Flash Point	Cool 4 deg C	PSEG01	K31	OK-COMP	04/11/2019	1010A
04/18/2019	Water	K2381-01	Flash Point	Cool 4 deg C	PSEG01	K11	CL-COMP-1	04/11/2019	1010A
04/18/2019	Water	K2381-02	Flash Point	Cool 4 deg C	PSEG01	K11	CL-COMP-2	04/11/2019	1010A
04/18/2019	Water	K2381-03	Flash Point	Cool 4 deg C	PSEG01	K11	CL-COMP-3	04/11/2019	1010A

Date/Time 04/12/19 13:40
 Received by: KS
 Relinquished by: CP

Date/Time 04/12/19 17:05
 Received by: CP
 Relinquished by: KS