



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:39
In Date: 04/12/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 04/13/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB102076

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2392-01	METE12	1	1.15	9.95	6.62	62.2
K2392-02	METE13	2	1.15	9.9	6.44	60.5
K2392-03	METE14	3	1.15	9.71	7.6	75.4
K2392-04	METE15	4	1.16	9.97	7.81	75.5
K2392-05	METE16	5	1.14	9.63	7.68	77.0
K2392-06	METE17	6	1.14	9.55	7.6	76.8
K2392-07	METE18	7	1.17	9.8	5.94	55.3
K2392-08	METE19	8	1.15	9.84	5.42	49.1
K2392-09	METE20	9	1.16	9.98	5.51	49.3
K2392-10	METE21	10	1.15	9.97	5.89	53.7
K2392-11	METE76	11	1.15	9.7	5.85	55.0
K2392-12	METE77	12	1.13	9.87	5.6	51.1
K2392-13	METE78	13	1.13	9.88	5.33	48.0
K2392-14	METE79	14	1.15	9.94	6.1	56.3
K2392-15	METE80	15	1.14	9.97	7.79	75.3
K2392-16	METE81	16	1.14	9.61	5.29	49.0
K2392-17	METE82	17	1.14	9.79	5.1	45.8
K2392-18	METE82D	18	1.14	9.79	5.1	45.8
K2392-19	METE82S	19	1.14	9.79	5.1	45.8
K2392-20	METE83	20	1.1	9.83	7.14	69.2
K2392-21	METE85	21	1.15	9.72	7.49	74.0
K2392-22	METE86	22	1.15	9.97	8.35	81.6

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$