



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/16/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 04/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:32
Out Date: 04/16/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB102080

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2395-01	METE57	1	1.16	9.79	6.22	58.6
K2395-02	METE58	2	1.16	9.86	5.25	47.0
K2395-03	METE59	3	1.15	9.95	6.1	56.3
K2395-04	METE60	4	1.19	9.71	6.21	58.9
K2395-05	METE61	5	1.11	9.94	6.59	62.1
K2395-06	METE62	6	1.14	9.68	7.35	72.7
K2395-07	METE63	7	1.17	9.98	8.42	82.3
K2395-08	METE64	8	1.12	9.62	7.54	75.5
K2395-09	METE74	9	1.16	9.79	6.28	59.3
K2395-10	METE84	10	1.13	9.94	5.6	50.7
K2395-11	METE65	11	1.16	9.94	6.36	59.2
K2395-12	METE66	12	1.17	9.92	6.2	57.5
K2395-13	METE67	13	1.13	9.52	5.05	46.7
K2395-14	METE68	14	1.16	9.6	5.38	50.0
K2395-15	METE69	15	1.18	9.97	6.04	55.3
K2395-16	METE70	16	1.17	9.98	5.62	50.5
K2395-17	METE71	17	1.19	9.82	5.67	51.9
K2395-18	METE72	18	1.16	9.97	5.91	53.9
K2395-19	METE73	19	1.17	9.71	5.94	55.9
K2395-20	METE87	20	1.14	9.87	7.88	77.2
K2395-21	METE87D	21	1.14	9.87	7.88	77.2
K2395-22	METE87S	22	1.14	9.87	7.88	77.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$