



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102092

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP73655
Residual chlorine std, 0PPM	WP73650
Residual chlorine std, 0.1PPM	WP73651
Residual chlorine LOQ, 0.2PPM	WP73652
Residual chlorine std, 0.2PPM	WP73653
Residual chlorine std, and CCV , 0.4PPM	WP73654
Residual chlorine std, 1.6PPM	WP73656
Residual chlorine ICV-LCS, 0.4PPM	WP73657
Residual chlorine LOD, 0.05PPM	WP73658
Total Chlorine Powder Pillows	W2249

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/15/2019	12:15
2	CAL2	0.1	1	0.000	0.100	0.100	04/15/2019	12:16
3	CAL3	0.2	1	0.000	0.200	0.200	04/15/2019	12:16
4	CAL4	0.4	1	0.000	0.400	0.400	04/15/2019	12:17
5	CAL5	0.8	1	0.000	0.800	0.800	04/15/2019	12:19
6	CAL6	1.6	1	0.000	1.600	1.600	04/15/2019	12:20



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102092

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3900	04/15/2019	12:21
2	ICB				1	0.0000	0.0000	0.0000	0.0000	04/15/2019	12:22
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4100	04/15/2019	12:22
4	CCB1				1	0.0000	0.0000	0.0000	0.0000	04/15/2019	12:23
5	LB102092BL	50	50		1	0.0000	0.0000	0.0000	0.0000	04/15/2019	12:24
6	LB102092BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.4000	04/15/2019	12:23
7	K2241-07	50	50		1	0.0000	0.0500	0.0500	0.0500	04/15/2019	12:24
8	K2241-08	50	50		1	0.0000	0.2100	0.2100	0.2100	04/15/2019	12:25
9	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4000	04/15/2019	12:27
10	CCB2				1	0.0000	0.0000	0.0000	0.0000	04/15/2019	12:28