



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102105

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/15/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	04/15/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.02	04/15/2019	17:12
4	ICV	1	Water	NA	NA	20.1	6.98	04/15/2019	17:13
5	CCV1	1	Water	NA	NA	20.1	2.01	04/15/2019	17:14
6	K2203-10	1	Solid	20.02	20	22.7	8.12	04/15/2019	17:15
7	K2203-12	1	Solid	20.03	20	22.3	7.97	04/15/2019	17:16
8	K2411-04	1	Solid	20.04	20	23.3	8.28	04/15/2019	17:17
9	K2411-08	1	Solid	20.02	20	23.0	7.86	04/15/2019	17:18
10	K2411-12	1	Solid	20.01	20	22.8	7.71	04/15/2019	17:19
11	K2421-01	1	Solid	20.03	20	22.5	7.22	04/15/2019	17:20
12	K2422-04	1	Solid	20.04	20	22.4	8.48	04/15/2019	17:21
13	K2422-08	1	Solid	20.03	20	22.1	8.38	04/15/2019	17:24
14	K2422-12	1	Solid	20.05	20	21.9	8.80	04/15/2019	17:25
15	K2422-12DUP	1	Solid	20.03	20	21.8	8.81	04/15/2019	17:26
16	CCV2	1	Water	NA	NA	20.1	12.02	04/15/2019	17:29

2B 102105

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2411

WorkList ID : 124842

Date : 04-15-2019 13:15:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2019	Solid	K2203-10	Corrosivity	Cool 4 deg C	PSEG05		SU-01-041519-A	04/15/2019	9045D
04/15/2019	Solid	K2203-12	Corrosivity	Cool 4 deg C	PSEG05		SU-01-041519-B	04/15/2019	9045D
04/13/2019	Solid	K2411-04	Corrosivity	Cool 4 deg C	PSEG01	K42	TP-16-S	04/13/2019	9045D
04/13/2019	Solid	K2411-08	Corrosivity	Cool 4 deg C	PSEG01	K42	TP-14-S	04/13/2019	9045D
04/13/2019	Solid	K2411-12	Corrosivity	Cool 4 deg C	PSEG01	K42	TP-14-D	04/13/2019	9045D
04/10/2019	Solid	K2421-01	Corrosivity	Cool 4 deg C	CLEA11	K11	OIL-WATER-SEPERATOR-1	04/10/2019	9045D
04/15/2019	Solid	K2422-04	Corrosivity	Cool 4 deg C	PSEG01	K52	TP1-S	04/15/2019	9045D
04/15/2019	Solid	K2422-08	Corrosivity	Cool 4 deg C	PSEG01	K52	TP1-D	04/15/2019	9045D
04/15/2019	Solid	K2422-12	Corrosivity	Cool 4 deg C	PSEG01	K52	TP3-S	04/15/2019	9045D

Date/Time 04-15-19 16:20
 Received by: Jn
 Relinquished by: Jn

Date/Time 04-15-19 18:10
 Received by: Jn
 Relinquished by: Jn