



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102106

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/15/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	04/15/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.02	04/15/2019	17:12
4	ICV	1	Water	NA	NA	20.1	6.98	04/15/2019	17:13
5	CCV1	1	Water	NA	NA	20.2	2.01	04/15/2019	17:35
6	K2414-01	1	Solid	20.02	20	21.5	6.06	04/15/2019	17:36
7	K2414-01DUP	1	Solid	20.03	20	21.6	6.07	04/15/2019	17:37
8	K2417-01	1	Solid	20.04	20	21.7	6.20	04/15/2019	17:38
9	CCV2	1	Water	NA	NA	20.2	12.02	04/15/2019	17:40

197102106

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k2414

WorkList ID : 124850

Date : 04-15-2019 16:45:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2019	Solid	K2414-01	pH	Cool 4 deg C	TULL02	K41	FK-PS-01-044	04/15/2019	9045D
04/15/2019	Solid	K2417-01	pH	Cool 4 deg C	TULL02	K31	FK-PS-01-045	04/15/2019	9045D

Date/Time 04-15-19 17:00
Received by: SN
Relinquished by: SN

Date/Time 04-15-19 18:20
Received by: SN
Relinquished by: SN