

LB102119


7471B

INSTRUMENT ID: CU1

Linear ▼

 μ Abs.:

61392



A= 0.0000e+000

B= 1.6280e-004

C= -3.3842e-002

Rho= 0.9999780

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	127	0.000	127				
0.2	0.200	0.189	-0.011	1371	0.0 %	1371				
2.5	2.500	2.505	0.005	15592	0.0 %	15592				
5.0	5.000	5.024	0.024	31066	0.0 %	31066				
7.5	7.500	7.535	0.035	46488	0.0 %	46488				
10.0	10.000	9.961	-0.039	61392	0.0 %	61392				

LB102119 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Apr 2019 11:30:01

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	127	-	0.00007471B	PPB		16 Apr 2019 12:25:46	S	Std
0.2 - 1	50-2	1371	-	0.20007471B	PPB		16 Apr 2019 12:28:05	S	Std
2.5 - 1	52-5	15592	-	2.50007471B	PPB		16 Apr 2019 12:30:23	S	Std
5.0 - 1	55-0	31066	-	5.00007471B	PPB		16 Apr 2019 12:32:40	S	Std
7.5 - 1	57-5	46488	-	7.50007471B	PPB		16 Apr 2019 12:34:56	S	Std
10.0 - 1	510-0	61392	-	10.00007471B	PPB		16 Apr 2019 12:37:13	S	Std
ICV01 - 1	ICV01	24533	3.9602	-7471B	PPB		16 Apr 2019 12:40:29	U	SMPL
ICB01 - 1	ICB01	10	-0.0322	-7471B	PPB		16 Apr 2019 12:42:44	U	SMPL
CCV05 - 1	CCV05	31203	5.0461	-7471B	PPB		16 Apr 2019 12:44:59	U	SMPL
CCB05 - 1	CCB05	32	-0.0286	-7471B	PPB		16 Apr 2019 12:47:14	U	SMPL
CRA - 1	CRA	1347	0.1855	-7471B	PPB		16 Apr 2019 12:49:30	U	SMPL
HighStd - 1	HighStd	62013	10.0620	-7471B	PPB		16 Apr 2019 12:51:44	U	SMPL
ChkStd - 1	ChkStd	42662	6.9116	-7471B	PPB		16 Apr 2019 12:54:00	U	SMPL
PB118933BL - 1	PBS	-33	-0.0392	-7471B	PPB		16 Apr 2019 12:56:18	U	SMPL
PB118933BS - 1	LCSS	20546	3.3111	-7471B	PPB		16 Apr 2019 12:58:35	U	SMPL
K2199-09 - 1	OR-02-041219-A	3895	0.6003	-7471B	PPB		16 Apr 2019 13:00:52	U	SMPL
K2199-09DUP - 1	OR-02-041219-ADUP	3916	0.6037	-7471B	PPB		16 Apr 2019 13:03:08	U	SMPL
K2199-09MS - 1	OR-02-041219-AMS	26466	4.2749	-7471B	PPB		16 Apr 2019 13:05:26	U	SMPL
K2199-09MSD - 1	OR-02-041219-AMSD	26397	4.2637	-7471B	PPB		16 Apr 2019 13:07:43	U	SMPL
K2199-11 - 1	OR-02-041219-B	5652	0.8863	-7471B	PPB		16 Apr 2019 13:09:58	U	SMPL
CCV06 - 1	CCV06	30871	4.9920	-7471B	PPB		16 Apr 2019 13:12:14	U	SMPL
CCB06 - 1	CCB06	30	-0.0290	-7471B	PPB		16 Apr 2019 13:14:30	U	SMPL
K2386-01 - 1	RBR-200048-49	4146	0.6411	-7471B	PPB		16 Apr 2019 13:16:46	U	SMPL
K2390-01 - 1	3425	1788	0.2572	-7471B	PPB		16 Apr 2019 13:19:02	U	SMPL
K2390-03 - 1	5124	1746	0.2504	-7471B	PPB		16 Apr 2019 13:21:18	U	SMPL
K2390-05 - 1	200050	1747	0.2506	-7471B	PPB		16 Apr 2019 13:23:34	U	SMPL
K2397-03 - 1	TP-1-S	247599	40.2759	-7471B	PPB		16 Apr 2019 13:25:50	U	SMPL
K2397-07 - 1	TP-1-D	242237	39.4029	-7471B	PPB		16 Apr 2019 13:28:06	U	SMPL
K2397-11 - 1	TP-3-S	314832	51.2216	-7471B	PPB		16 Apr 2019 13:30:25	U	SMPL
K2397-15 - 1	TP-3-D	190286	30.9452	-7471B	PPB		16 Apr 2019 13:32:45	U	SMPL
K2405-02 - 1	B-1D(0-24)	1407	0.1952	-7471B	PPB		16 Apr 2019 13:35:05	U	SMPL
K2407-03 - 1	TP-11-S	12233	1.9577	-7471B	PPB		16 Apr 2019 13:37:24	U	SMPL
CCV07 - 1	CCV07	30960	5.0065	-7471B	PPB		16 Apr 2019 13:39:40	U	SMPL
CCB07 - 1	CCB07	-15	-0.0363	-7471B	PPB		16 Apr 2019 13:41:56	U	SMPL
K2407-07 - 1	TP-11-D	3837	0.5908	-7471B	PPB		16 Apr 2019 13:44:12	U	SMPL
K2407-11 - 1	TP-17-S	5505	0.8624	-7471B	PPB		16 Apr 2019 13:46:27	U	SMPL
K2411-03 - 1	TP-16-S	4518	0.7017	-7471B	PPB		16 Apr 2019 13:48:43	U	SMPL
K2411-07 - 1	TP-14-S	3412	0.5216	-7471B	PPB		16 Apr 2019 13:50:59	U	SMPL
K2411-11 - 1	TP-14-D	13149	2.1068	-7471B	PPB		16 Apr 2019 13:53:15	U	SMPL
K2413-01 - 1	NCC-TOPSOIL-KY	4462	0.6926	-7471B	PPB		16 Apr 2019 13:55:31	U	SMPL
K2413-02 - 1	NCC-TOPSOIL-CR	3809	0.5863	-7471B	PPB		16 Apr 2019 13:57:48	U	SMPL
K2414-01 - 1	FK-PS-01-044	2630	0.3943	-7471B	PPB		16 Apr 2019 14:00:05	U	SMPL
PB118934BL - 1	PBS	55	-0.0249	-7471B	PPB		16 Apr 2019 14:02:22	U	SMPL
PB118934BS - 1	LCSS	20498	3.3033	-7471B	PPB		16 Apr 2019 14:04:38	U	SMPL
CCV08 - 1	CCV08	29905	4.8348	-7471B	PPB		16 Apr 2019 14:06:54	U	SMPL
CCB08 - 1	CCB08	37	-0.0278	-7471B	PPB		16 Apr 2019 14:09:10	U	SMPL
K2203-09 - 1	SU-01-041519-A	8675	1.3785	-7471B	PPB		16 Apr 2019 14:11:26	U	SMPL
K2203-11 - 1	SU-01-041519-B	2966	0.4490	-7471B	PPB		16 Apr 2019 14:13:42	U	SMPL
K2389-02 - 1	STONE-1	195	-0.0021	-7471B	PPB		16 Apr 2019 14:15:58	U	SMPL
K2389-03 - 1	STONE-2	207	-0.0001	-7471B	PPB		16 Apr 2019 14:18:14	U	SMPL
K2389-04 - 1	STONE-3	256	0.0078	-7471B	PPB		16 Apr 2019 14:20:30	U	SMPL
K2389-05 - 1	STONE-4	210	0.0003	-7471B	PPB		16 Apr 2019 14:22:46	U	SMPL
K2417-01 - 1	FK-PS-01-045	3578	0.5487	-7471B	PPB		16 Apr 2019 14:25:03	U	SMPL
K2422-03 - 1	TP1-S	5945	0.9340	-7471B	PPB		16 Apr 2019 14:27:20	U	SMPL
K2422-07 - 1	TP1-D	6984	1.1032	-7471B	PPB		16 Apr 2019 14:29:37	U	SMPL
K2422-11 - 1	TP3-S	6962	1.0996	-7471B	PPB		16 Apr 2019 14:31:53	U	SMPL
CCV09 - 1	CCV09	29943	4.8410	-7471B	PPB		16 Apr 2019 14:34:09	U	SMPL
CCB09 - 1	CCB09	-7	-0.0350	-7471B	PPB		16 Apr 2019 14:36:25	U	SMPL
K2423-01 - 1	TP1-A	1114	0.1475	-7471B	PPB		16 Apr 2019 14:38:41	U	SMPL
K2423-03 - 1	TP1-B	841	0.1031	-7471B	PPB		16 Apr 2019 14:40:57	U	SMPL
K2423-05 - 1	TP2-A	2457	0.3662	-7471B	PPB		16 Apr 2019 14:43:13	U	SMPL
K2423-07 - 1	TP2-B	1074	0.1410	-7471B	PPB		16 Apr 2019 14:45:29	U	SMPL
K2423-09 - 1	TP3-A	1378	0.1905	-7471B	PPB		16 Apr 2019 14:47:45	U	SMPL
K2423-11 - 1	TP3-B	624	0.0677	-7471B	PPB		16 Apr 2019 14:50:02	U	SMPL
K2423-13 - 1	TP4-A	1293	0.1767	-7471B	PPB		16 Apr 2019 14:52:18	U	SMPL
K2423-15 - 1	TP4-B	1425	0.1982	-7471B	PPB		16 Apr 2019 14:54:35	U	SMPL
K2423-17 - 1	TP5-A	2094	0.3071	-7471B	PPB		16 Apr 2019 14:56:52	U	SMPL
K2423-19 - 1	TP5-B	2554	0.3820	-7471B	PPB		16 Apr 2019 14:59:09	U	SMPL
CCV10 - 1	CCV10	30277	4.8953	-7471B	PPB		16 Apr 2019 15:01:26	U	SMPL
CCB10 - 1	CCB10	9	-0.0324	-7471B	PPB		16 Apr 2019 15:03:42	U	SMPL
K2423-19DUP - 1	TP5-BDUP	2514	0.3754	-7471B	PPB		16 Apr 2019 15:05:58	U	SMPL
K2423-19MS - 1	TP5-BMS	25097	4.0520	-7471B	PPB		16 Apr 2019 15:08:15	U	SMPL
K2423-19MSD - 1	TP5-BMSD	24742	3.9942	-7471B	PPB		16 Apr 2019 15:10:31	U	SMPL
K2199-09LX5 - 1		759	0.0897	-7471B	PPB		16 Apr 2019 15:12:48	U	SMPL
K2199-09A - 1		26544	4.2876	-7471B	PPB		16 Apr 2019 15:15:04	U	SMPL
K2423-19LX5 - 1		501	0.0477	-7471B	PPB		16 Apr 2019 15:17:21	U	SMPL
K2423-19A - 1		25067	4.0471	-7471B	PPB		16 Apr 2019 15:19:37	U	SMPL
K2397-03DLX10 - 1		24828	4.0082	-7471B	PPB		16 Apr 2019 15:21:54	U	SMPL
K2397-07DLX10 - 1		25260	4.0785	-7471B	PPB		16 Apr 2019 15:24:11	U	SMPL
K2397-11DLX10 - 1		33282	5.3846	-7471B	PPB		16 Apr 2019 15:26:28	U	SMPL
CCV11 - 1	CCV11	30210	4.8844	-7471B	PPB		16 Apr 2019 15:28:44	U	SMPL

LB102119 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Apr 2019 11:30:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB11 - 1	CCB11	18	-0.0309	-7471B	PPB		16 Apr 2019 15:31:02	U	SMPL
K2397-15DLX10 - 1		18678	3.0070	-7471B	PPB		16 Apr 2019 15:33:18	U	SMPL
CCV12 - 1	CCV12	29547	4.7765	-7471B	PPB		16 Apr 2019 15:35:34	U	SMPL
CCB12 - 1	CCB12	11	-0.0321	-7471B	PPB		16 Apr 2019 15:37:51	U	SMPL