



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102122

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/16/2019	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/16/2019	11:11
3	CAL3	1	Water	NA	NA	20.2	10.02	04/16/2019	11:12
4	ICV	1	Water	NA	NA	20.1	6.98	04/16/2019	11:13
5	CCV1	1	Water	NA	NA	20.1	2.02	04/16/2019	11:15
6	K2427-01	1	Water	NA	NA	16.1	5.90	04/16/2019	11:16
7	K2427-01DUP	1	Water	NA	NA	16.2	5.91	04/16/2019	11:18
8	K2427-02	1	Water	NA	NA	17.2	4.33	04/16/2019	11:20
9	K2427-03	1	Water	NA	NA	17.7	4.90	04/16/2019	11:21
10	CCV2	1	Water	NA	NA	20.1	12.02	04/16/2019	11:24

2B102122

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2427

WorkList ID : 124896

Date : 04-16-2019 10:58:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2019	Water	K2427-01-1A	pH	Cool 4 deg C	STER09	K21	CELL-A-POOL-WATER(DE-L)	04/15/2019	SM 4500-H B
04/15/2019	Water	K2427-02-1A	pH	Cool 4 deg C	STER09	K21	CELL-B-C-POOL-WATER(DE-L)	04/15/2019	SM 4500-H B
04/15/2019	Water	K2427-03-1A	pH	Cool 4 deg C	STER09	K21	CELL-D-POOL-WATER(DE-L)	04/15/2019	SM 4500-H B

Date/Time 04-16-19 11:00
 Received by: JD
 Relinquished by: CO

Date/Time 04-16-19 12:05
 Received by: cm
 Relinquished by: JD