



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB102123

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP73610
COD calibration std. 100 ppm	WP73607
COD calibration std. 10 ppm	WP73614
COD calibration std. 150 ppm	WP73609
COD calibration std. 50 ppm	WP73608
COD calibration std. 0 ppm	WP73605
COD Digestion Vials Low Level 0-150Mg/L	W2470
COD ICV-LCS std, 50ppm	WP73752
COD CCV std, 50ppm	WP73753
COD LOD std, 5ppm	WP73754

Temp In (C): <u>150</u>	Date In: <u>04/16/2019</u>	Time In: <u>10:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>04/16/2019</u>	Time Out: <u>12:01</u>

Intercept: -0.3388

Slope: 1.0087

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	04/12/2019	12:40
2	CAL2	10	1	Water	10.000	04/12/2019	12:40
3	CAL3	50	1	Water	49.000	04/12/2019	12:41
4	CAL4	100	1	Water	101.000	04/12/2019	12:41
5	CAL5	150	1	Water	151.000	04/12/2019	12:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB102123

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.905	04/12/2019	12:42
2	ICB				1	Water	0.000	0.336	04/12/2019	12:43
3	CCV1	50			1	Water	50.000	49.905	04/16/2019	13:00
4	CCB1				1	Water	0.000	0.336	04/16/2019	13:00
5	LB102123BL				1	Water	0.000	0.336	04/16/2019	13:01
6	LB102123BS	50			1	Water	49.000	48.913	04/16/2019	13:01
7	K2241-09				1	Water	4.000	4.301	04/16/2019	13:02
8	K2397-17				1	Water	33.000	33.051	04/16/2019	13:02
9	K2424-01				1	Water	10.000	10.250	04/16/2019	13:03
10	K2424-01DUP				1	Water	11.000	11.241	04/16/2019	13:03
11	K2424-01MS	50			1	Water	64.000	63.784	04/16/2019	13:04
12	K2424-01MSD	50			1	Water	63.000	62.793	04/16/2019	13:04
13	CCV2	50			1	Water	50.000	49.905	04/16/2019	13:05
14	CCB2				1	Water	0.000	0.336	04/16/2019	13:05