



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102129

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP73736
Residual chlorine std, 0PPM	WP73730
Residual chlorine std, 0.1PPM	WP73732
Residual chlorine std, 0.2PPM	WP73734
Residual chlorine std, and CCV , 0.4PPM	WP73735
Residual chlorine std, 1.6PPM	WP73737
Residual chlorine ICV-LCS, 0.4PPM	WP73738
Residual chlorine LOD, 0.05PPM	WP73731
Total Chlorine Powder Pillows	W2249

Intercept: -0.0014

Slope: 1.0060

Regression: 0.999995

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/16/2019	13:40
2	CAL2	0.1	1	0.000	0.100	0.100	04/16/2019	13:42
3	CAL3	0.2	1	0.000	0.200	0.200	04/16/2019	13:43
4	CAL4	0.4	1	0.000	0.400	0.400	04/16/2019	13:45
5	CAL5	0.8	1	0.000	0.800	0.800	04/16/2019	13:46
6	CAL6	1.6	1	0.000	1.610	1.610	04/16/2019	13:47



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102129

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3900	04/16/2019	13:29
2	ICB				1	0.0000	0.0000	0.0000	0.0000	04/16/2019	13:30
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.4000	04/16/2019	13:50
4	CCB1				1	0.0000	0.0000	0.0000	0.0000	04/16/2019	13:51
5	LB102129BL	50	50		1	0.0000	0.0000	0.0000	0.0000	04/16/2019	13:55
6	LB102129BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3900	04/16/2019	13:56
7	K2241-09	50	50		1	0.0000	0.0400	0.0400	0.0400	04/16/2019	13:57
8	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4000	04/16/2019	13:58
9	CCB2				1	0.0000	0.0000	0.0000	0.0000	04/16/2019	13:59