

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/16/2019 15:33

Test: Phenolics-

Sample Id	Result ^{mg/d}	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.123	
ICB1	-0.001	0.0	0.013	
CCV1	0.997	0.0	0.122	
CCB1	0.002	0.0	0.013	
PB118870BL	-0.005	0.0	0.013	
PB118870BS	1.013	0.0	0.123	
K2241-03	0.008	0.0	0.014	
PB118871BL	-0.003	0.0	0.013	
PB118871BS	1.010	0.0	0.123	
K2241-09	0.010	0.0	0.014	
CCV2	0.998	0.0	0.122	
CCB2	-0.003	0.0	0.013	

N	12
Mean	0.419
SD	0.5170
CV%	123.27

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 16 14:57:00 2019

Tue Apr 16 15:29:59 2019

AS

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0454	mg/l	4/16/2019 14:57:00	
0.1PPM	A	Phenolics-	P	0.0995	mg/l	4/16/2019 14:57:01	
0.25PPM	A	Phenolics-	P	0.2504	mg/l	4/16/2019 14:57:02	
0.50PPM	A	Phenolics-	P	0.5023	mg/l	4/16/2019 14:57:03	
1.0PPM	A	Phenolics-	P	1.0059	mg/l	4/16/2019 14:57:04	
2.0PPM	A	Phenolics-	P	1.9966	mg/l	4/16/2019 14:57:05	
ICV1	S	Phenolics-	P	1.0089	mg/l	4/16/2019 15:29:48	
ICB1	S	Phenolics-	P	-0.0005	mg/l	4/16/2019 15:29:49	
CCV1	S	Phenolics-	P	0.9966	mg/l	4/16/2019 15:29:50	
CCB1	S	Phenolics-	P	0.0019	mg/l	4/16/2019 15:29:51	
PB118870BL	S	Phenolics-	P	-0.0054	mg/l	4/16/2019 15:29:52	
PB118870BS	S	Phenolics-	P	1.0125	mg/l	4/16/2019 15:29:53	
K2241-03	S	Phenolics-	P	0.0078	mg/l	4/16/2019 15:29:54	
PB118871BL	S	Phenolics-	P	-0.0031	mg/l	4/16/2019 15:29:55	
PB118871BS	S	Phenolics-	P	1.0099	mg/l	4/16/2019 15:29:56	
K2241-09	S	Phenolics-	P	0.0103	mg/l	4/16/2019 15:29:57	
CCV2	S	Phenolics-	P	0.9976	mg/l	4/16/2019 15:29:58	
CCB2	S	Phenolics-	P	-0.0035	mg/l	4/16/2019 15:29:59	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/16/2019 15:08


Test Phenolics-

Accepted 4/16/2019 15:07

Factor 9.205
Bias 0.013

Coeff. of det. 0.999974

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.018	0.0454	0.0500	
2	Phenol-2	0.024	0.0995	0.1000	
3	Phenol-2	0.040	0.2504	0.2500	
4	Phenol-2	0.068	0.5023	0.5000	
5	Phenol-2	0.123	1.0059	1.0000	
6	Phenol-2	0.230	1.9966	2.0000	