



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102141

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/16/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.01	04/16/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.03	04/16/2019	18:03
4	ICV	1	Water	NA	NA	20.1	7.00	04/16/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.02	04/16/2019	18:35
6	K2437-01	1	Solid	20.03	20	21.7	5.92	04/16/2019	18:36
7	K2437-01DUP	1	Solid	20.04	20	21.8	5.93	04/16/2019	18:37
8	CCV2	1	Water	NA	NA	20.1	12.01	04/16/2019	18:40

WORKLIST(Hardcopy Internal Chain)

187102141

WorkList Name : PJH K2437

WorkList ID : 124912

Date : 04-16-2019 17:17:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/16/2019	Solid	K2437-01	pH	Cool 4 deg C	TULL02	K22	FK-PS-01-046	04/16/2019	9045D

Date/Time 04-16-19 17:20
 Received by: jm
 Relinquished by: cm

Date/Time 04-16-19 18:30
 Received by: cm
 Relinquished by: jm