

LB102/47


7470A

INSTRUMENT ID: C01

Linear

 μ Abs.:

42481



A= 0.0000e+000

B= 2.3670e-004

C= -1.1289e-002

Rho= 0.9997743

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	60	0.000	60				
0.2	0.200	0.198	-0.002	884	0.0 %	884				
2.5	2.500	2.466	-0.034	10465	0.0 %	10465				
5.0	5.000	5.124	0.124	21697	0.0 %	21697				
7.5	7.500	7.365	-0.135	31162	0.0 %	31162				
10.0	10.000	10.044	0.044	42481	0.0 %	42481				

LB102147 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Apr 2019 10:17:32

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	60	-	0.00007470A	PPB		17 Apr 2019 10:31:16	S	Std
0.2 - 1	50.2	884	-	0.20007470A	PPB		17 Apr 2019 10:33:24	S	Std
2.5 - 1	52.5	10465	-	2.50007470A	PPB		17 Apr 2019 10:35:33	S	Std
5.0 - 1	55.0	21697	-	5.00007470A	PPB		17 Apr 2019 10:37:41	S	Std
7.5 - 1	57.5	31162	-	7.50007470A	PPB		17 Apr 2019 10:39:48	S	Std
10.0 - 1	510.0	42481	-	10.00007470A	PPB		17 Apr 2019 10:41:55	S	Std
ICV03 - 1	ICV03	16862	3.9800	-7470A	PPB		17 Apr 2019 10:44:44	U	SMPL
ICB03 - 1	ICB03	9	-0.0092	-7470A	PPB		17 Apr 2019 10:46:50	U	SMPL
CCV17 - 1	CCV17	21791	5.1467	-7470A	PPB		17 Apr 2019 10:48:57	U	SMPL
CCB17 - 1	CCB17	22	-0.0061	-7470A	PPB		17 Apr 2019 10:51:02	U	SMPL
CRA - 1	CRA	912	0.2046	-7470A	PPB		17 Apr 2019 10:53:08	U	SMPL
HighStd - 1	HighStd	42851	10.1316	-7470A	PPB		17 Apr 2019 10:55:14	U	SMPL
ChkStd - 1	ChkStd	29506	6.9728	-7470A	PPB		17 Apr 2019 10:57:21	U	SMPL
PB118976BL - 1	PBW	-15	-0.0148	-7470A	PPB		17 Apr 2019 10:59:28	U	SMPL
PB118976BS - 1	LCSW	15272	3.6036	-7470A	PPB		17 Apr 2019 11:01:34	U	SMPL
K2203-10 - 1	SU-01-041519-A	17	-0.0073	-7470A	PPB		17 Apr 2019 11:03:42	U	SMPL
K2203-10DUP - 1	SU-01-041519-ADUP	69	0.0050	-7470A	PPB		17 Apr 2019 11:05:49	U	SMPL
K2203-10MS - 1	SU-01-041519-AMS	16377	3.8652	-7470A	PPB		17 Apr 2019 11:07:57	U	SMPL
K2203-10MSD - 1	SU-01-041519-AMSD	16082	3.7954	-7470A	PPB		17 Apr 2019 11:10:04	U	SMPL
K2203-12 - 1	SU-01-041519-B	26	-0.0051	-7470A	PPB		17 Apr 2019 11:12:10	U	SMPL
CCV18 - 1	CCV18	21141	4.9928	-7470A	PPB		17 Apr 2019 11:14:16	U	SMPL
CCB18 - 1	CCB18	86	0.0091	-7470A	PPB		17 Apr 2019 11:16:23	U	SMPL
K2411-04 - 1	TP-16-S	53	0.0013	-7470A	PPB		17 Apr 2019 11:18:28	U	SMPL
K2411-08 - 1	TP-14-S	34	-0.0032	-7470A	PPB		17 Apr 2019 11:20:34	U	SMPL
K2411-12 - 1	TP-14-D	49	0.0003	-7470A	PPB		17 Apr 2019 11:22:40	U	SMPL
K2421-01 - 1	OIL-WATER-SEPERATOR-1	141	0.0221	-7470A	PPB		17 Apr 2019 11:24:47	U	SMPL
K2422-04 - 1	TP1-S	52	0.0010	-7470A	PPB		17 Apr 2019 11:26:54	U	SMPL
K2422-08 - 1	TP1-D	45	-0.0006	-7470A	PPB		17 Apr 2019 11:29:01	U	SMPL
K2422-12 - 1	TP3-S	39	-0.0021	-7470A	PPB		17 Apr 2019 11:31:09	U	SMPL
PB118977BL - 1	PBW	27	-0.0049	-7470A	PPB		17 Apr 2019 11:33:17	U	SMPL
PB118977BS - 1	LCSW	15753	3.7175	-7470A	PPB		17 Apr 2019 11:35:25	U	SMPL
K2397-18 - 1	TP-3	-18	-0.0155	-7470A	PPB		17 Apr 2019 11:37:32	U	SMPL
CCV19 - 1	CCV19	21589	5.0989	-7470A	PPB		17 Apr 2019 11:39:38	U	SMPL
CCB19 - 1	CCB19	-7	-0.0129	-7470A	PPB		17 Apr 2019 11:41:45	U	SMPL
K2397-18DUP - 1	TP-3DUP	18	-0.0070	-7470A	PPB		17 Apr 2019 11:43:51	U	SMPL
K2397-18MS - 1	TP-3MS	16827	3.9717	-7470A	PPB		17 Apr 2019 11:45:58	U	SMPL
K2397-18MSD - 1	TP-3MSD	16538	3.9033	-7470A	PPB		17 Apr 2019 11:48:04	U	SMPL
PB118983BL - 1	PBW	-9	-0.0134	-7470A	PPB		17 Apr 2019 11:50:11	U	SMPL
PB118983BS - 1	LCSW	15363	3.6252	-7470A	PPB		17 Apr 2019 11:52:17	U	SMPL
K2397-17 - 1	TP-3	28551	6.7468	-7470A	PPB		17 Apr 2019 11:54:24	U	SMPL
K2397-17DUP - 1	TP-3DUP	28524	6.7404	-7470A	PPB		17 Apr 2019 11:56:31	U	SMPL
K2397-17MS - 1	TP-3MS	45792	10.8278	-7470A	PPB		17 Apr 2019 11:58:39	U	SMPL
K2397-17MSD - 1	TP-3MSD	45787	10.8266	-7470A	PPB		17 Apr 2019 12:00:47	U	SMPL
PB118984BL - 1	PBW	-46	-0.0222	-7470A	PPB		17 Apr 2019 12:02:53	U	SMPL
CCV20 - 1	CCV20	21339	5.0397	-7470A	PPB		17 Apr 2019 12:05:00	U	SMPL
CCB20 - 1	CCB20	-9	-0.0134	-7470A	PPB		17 Apr 2019 12:07:07	U	SMPL
PB118984BS - 1	LCSW	15846	3.7395	-7470A	PPB		17 Apr 2019 12:09:13	U	SMPL
K2241-07 - 1	LOD-MDL-WATER-01-QT2-2019	315	0.0633	-7470A	PPB		17 Apr 2019 12:11:20	U	SMPL
K2241-08 - 1	LOQ-WATER-02-QT2-2019	1870	0.4313	-7470A	PPB		17 Apr 2019 12:16:08	U	SMPL
PB118976TB - 1	PB118899TB	54	0.0015	-7470A	PPB		17 Apr 2019 12:18:14	U	SMPL
PB118977TB - 1	PB118901TB	60	0.0029	-7470A	PPB		17 Apr 2019 12:20:20	U	SMPL
K2203-10LX5 - 1		8	-0.0094	-7470A	PPB		17 Apr 2019 12:22:27	U	SMPL
K2203-10A - 1		16540	3.9038	-7470A	PPB		17 Apr 2019 12:24:35	U	SMPL
K2397-18LX5 - 1		-43	-0.0215	-7470A	PPB		17 Apr 2019 12:26:43	U	SMPL
K2397-18A - 1		16608	3.9199	-7470A	PPB		17 Apr 2019 12:28:51	U	SMPL
K2397-17LX5 - 1		4884	1.1448	-7470A	PPB		17 Apr 2019 12:30:58	U	SMPL
CCV21 - 1	CCV21	21522	5.0830	-7470A	PPB		17 Apr 2019 12:33:05	U	SMPL
CCB21 - 1	CCB21	-32	-0.0189	-7470A	PPB		17 Apr 2019 12:35:11	U	SMPL
K2397-17A - 1		45843	10.8399	-7470A	PPB		17 Apr 2019 12:37:18	U	SMPL
K2397-17DLX2 - 1		13965	3.2943	-7470A	PPB		17 Apr 2019 12:39:24	U	SMPL
K2397-17DUPDLX2 - 1		14040	3.3120	-7470A	PPB		17 Apr 2019 12:41:31	U	SMPL
K2397-17MSDLX2 - 1		22870	5.4021	-7470A	PPB		17 Apr 2019 12:43:38	U	SMPL
K2397-17MSDDLX2 - 1		22664	5.3533	-7470A	PPB		17 Apr 2019 12:45:44	U	SMPL
K2397-17ADLX2 - 1		22195	5.2423	-7470A	PPB		17 Apr 2019 12:47:52	U	SMPL
CCV22 - 1	CCV22	21601	5.1017	-7470A	PPB		17 Apr 2019 12:49:59	U	SMPL
CCB22 - 1	CCB22	3	-0.0106	-7470A	PPB		17 Apr 2019 12:52:07	U	SMPL