

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB102151

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP73774
calibration std. phosphate 0.5 ppm	WP73773
calibration std. phosphate 0.3 ppm	WP73772
calibration std. 0 ppm	WP73769
phosphate ICV std.	WP73776
phosphate CCV std.	WP73775
calibration std. phosphate 0.1 ppm	WP73771
calibration std. phosphate 0.05 ppm	WP73770
Combined reagent	WP73795
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0008Slope: 0.6508Regression: 0.999843

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/17/2019	11:18
2	CAL2	0.05	1	50	50	0.033	04/17/2019	11:18
3	CAL3	0.10	1	50	50	0.066	04/17/2019	11:19
4	CAL4	0.30	1	50	50	0.186	04/17/2019	11:20
5	CAL5	0.50	1	50	50	0.329	04/17/2019	11:20
6	CAL6	1.00	1	50	50	0.650	04/17/2019	11:21



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB102151

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.490	04/17/2019	11:22
2	ICB		1	50	50	0.000	0.001	04/17/2019	11:23
3	CCV1	0.50	1	50	50	0.326	0.502	04/17/2019	11:24
4	CCB1		1	50	50	0.000	0.001	04/17/2019	11:25
5	PB118970BL		1	50	50	0.000	0.001	04/17/2019	11:26
6	PB118970BS	0.50	1	50	50	0.302	0.465	04/17/2019	11:26
7	K2424-01		1	50	50	0.004	0.007	04/17/2019	11:27
8	K2424-01DUP		1	50	50	0.005	0.009	04/17/2019	11:27
9	K2424-01MS	0.50	1	50	50	0.321	0.494	04/17/2019	11:28
10	K2424-01MSD	0.50	1	50	50	0.319	0.491	04/17/2019	11:29
11	CCV2	0.50	1	50	50	0.325	0.501	04/17/2019	11:29
12	CCB2		1	50	50	0.000	0.001	04/17/2019	11:30