



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102186

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/18/2019	10:10
2	CAL2	1	Water	NA	NA	20.1	7.01	04/18/2019	10:11
3	CAL3	1	Water	NA	NA	20.1	10.03	04/18/2019	10:12
4	ICV	1	Water	NA	NA	20.2	6.99	04/18/2019	10:15
5	CCV1	1	Water	NA	NA	20.1	2.01	04/18/2019	10:16
6	K2461-01	1	Water	NA	NA	12.2	6.31	04/18/2019	10:17
7	K2461-01DUP	1	Water	NA	NA	12.1	6.32	04/18/2019	10:18
8	K2461-02	1	Water	NA	NA	13.1	6.46	04/18/2019	10:19
9	K2461-03	1	Water	NA	NA	12.5	6.46	04/18/2019	10:21
10	K2461-04	1	Water	NA	NA	12.1	7.87	04/18/2019	10:23
11	K2461-05	1	Water	NA	NA	12.6	9.62	04/18/2019	10:25
12	K2461-06	1	Water	NA	NA	13.2	6.62	04/18/2019	10:26
13	CCV2	1	Water	NA	NA	20.1	12.01	04/18/2019	10:27

LB 102186

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2461

WorkList ID : 125008

Date : 04-18-2019 09:51:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/17/2019	Water	K2461-01 - K	pH	Cool 4 deg C	WOOD06	K62	PZ-01	04/17/2019	9040C
04/17/2019	Water	K2461-02 - K	pH	Cool 4 deg C	WOOD06	K62	PZ-02	04/17/2019	9040C
04/17/2019	Water	K2461-03 - K	pH	Cool 4 deg C	WOOD06	K62	PZ-15	04/17/2019	9040C
04/17/2019	Water	K2461-04 - K	pH	Cool 4 deg C	WOOD06	K62	GMW-5	04/17/2019	9040C
04/17/2019	Water	K2461-05 - K	pH	Cool 4 deg C	WOOD06	K62	GMW-11	04/17/2019	9040C
04/17/2019	Water	K2461-06 - K	pH	Cool 4 deg C	WOOD06	K62	GMW-1	04/17/2019	9040C

Date/Time 04-18-19 10:00
Received by: SA
Relinquished by: CS

Date/Time 04-18-19 11:15
Received by: CS
Relinquished by: SA