

LB102198


245.1

INSTRUMENT ID : CU1

Linear

 μ Abs.:

44392



A= 0.0000e+000

B= 2.2529e-004

C= -8.5479e-002

Rho= 0.9991026

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	339	0.000	339				
0.05	0.050	0.051	0.001	606	0.0 %	606				
0.2	0.200	0.178	-0.022	1168	0.0 %	1168				
2.5	2.500	2.355	-0.145	10831	0.0 %	10831				
5.0	5.000	5.365	0.365	24193	0.0 %	24193				
7.5	7.500	7.395	-0.105	33204	0.0 %	33204				
10.0	10.000	9.916	-0.084	44392	0.0 %	44392				

LB102198 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Apr 2019 13:57:31

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	339	-	0.0000	245.1	PPB	18 Apr 2019 13:58:29	S	Std
0.05 - 1	50-05	606	-	0.0500	245.1	PPB	18 Apr 2019 14:00:46	S	Std
0.2 - 1	50-2	1168	-	0.2000	245.1	PPB	18 Apr 2019 14:03:04	S	Std
2.5 - 1	52-5	10831	-	2.5000	245.1	PPB	18 Apr 2019 14:05:21	S	Std
5.0 - 1	55-0	24193	-	5.0000	245.1	PPB	18 Apr 2019 14:07:38	S	Std
7.5 - 1	57-5	33204	-	7.5000	245.1	PPB	18 Apr 2019 14:09:54	S	Std
10.0 - 1	510-0	44392	-	10.0000	245.1	PPB	18 Apr 2019 14:12:10	S	Std
ICV07 - 1	ICV07	18335	4.0452	-245.1		PPB	18 Apr 2019 14:18:55	U	SMPL
ICB07 - 1	ICB07	279	-0.0226	-245.1		PPB	18 Apr 2019 14:21:10	U	SMPL
CCV36 - 1	CCV36	24261	5.3803	-245.1		PPB	18 Apr 2019 14:23:25	U	SMPL
CCB36 - 1	CCB36	264	-0.0260	-245.1		PPB	18 Apr 2019 14:25:40	U	SMPL
CRA - 1	CRA	600	0.0497	-245.1		PPB	18 Apr 2019 14:27:56	U	SMPL
HighStd - 1	HighStd	44560	9.9535	-245.1		PPB	18 Apr 2019 14:30:12	U	SMPL
ChkStd - 1	ChkStd	30264	6.7327	-245.1		PPB	18 Apr 2019 14:32:27	U	SMPL
PB119024BL - 1	PBW	208	-0.0386	-245.1		PPB	18 Apr 2019 14:34:44	U	SMPL
PB119024BS - 1	LCSW	17137	3.7753	-245.1		PPB	18 Apr 2019 14:36:59	U	SMPL
K2408-01 - 1	EFF-WW	93	-0.0645	-245.1		PPB	18 Apr 2019 14:39:15	U	SMPL
K2408-01DUP - 1	EFF-WWDUP	147	-0.0524	-245.1		PPB	18 Apr 2019 14:41:33	U	SMPL
K2408-01MS - 1	EFF-WWMS	18964	4.1869	-245.1		PPB	18 Apr 2019 14:47:30	U	SMPL
K2408-01MSD - 1	EFF-WWMSD	18545	4.0925	-245.1		PPB	18 Apr 2019 14:49:47	U	SMPL
K2446-01 - 1	COMP	352	-0.0062	-245.1		PPB	18 Apr 2019 14:52:03	U	SMPL
CCV37 - 1	CCV37	21001	4.6459	-245.1		PPB	18 Apr 2019 14:54:19	U	SMPL
CCB37 - 1	CCB37	236	-0.0323	-245.1		PPB	18 Apr 2019 14:56:34	U	SMPL
PB119026BL - 1	PBW	304	-0.0170	-245.1		PPB	18 Apr 2019 14:58:49	U	SMPL
PB119026BS - 1	LCSW	16576	3.6490	-245.1		PPB	18 Apr 2019 15:01:04	U	SMPL
K2241-10 - 1	LOD-MDL-WATER-04-QT2-2019	573	0.0436	-245.1		PPB	18 Apr 2019 15:03:20	U	SMPL
K2241-11 - 1	LOQ-WATER-05-QT2-2019	2379	0.4505	-245.1		PPB	18 Apr 2019 15:15:39	U	SMPL
K2408-01LX5 - 1		1	-0.0853	-245.1		PPB	18 Apr 2019 15:17:55	U	SMPL
K2408-01A - 1		19630	4.3370	-245.1		PPB	18 Apr 2019 15:23:45	U	SMPL
CCV38 - 1	CCV38	22228	4.9223	-245.1		PPB	18 Apr 2019 15:26:02	U	SMPL
CCB38 - 1	CCB38	223	-0.0352	-245.1		PPB	18 Apr 2019 15:28:19	U	SMPL