

LB102205


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs.:

43521



A= 0.0000e+000

B= 2.2937e-004

C= -6.1849e-002

Rho= 0.9997404

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	418	0.000	418				
0.2	0.200	0.203	0.003	1153	0.0 %	1153				
2.5	2.500	2.450	-0.050	10950	0.0 %	10950				
5.0	5.000	4.928	-0.072	21756	0.0 %	21756				
7.5	7.500	7.665	0.165	33688	0.0 %	33688				
10.0	10.000	9.920	-0.080	43521	0.0 %	43521				

LB102205 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Apr 2019 15:48:18

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	418	-	0.00007470A	PPB	18 Apr 2019 15:49:24	S	Std
0.2 - 1	50.2	1153	-	0.20007470A	PPB	18 Apr 2019 15:51:32	S	Std
2.5 - 1	52.5	10950	-	2.50007470A	PPB	18 Apr 2019 15:53:40	S	Std
5.0 - 1	55.0	21756	-	5.00007470A	PPB	18 Apr 2019 15:55:48	S	Std
7.5 - 1	57.5	33688	-	7.50007470A	PPB	18 Apr 2019 15:57:55	S	Std
10.0 - 1	510.0	43521	-	10.00007470A	PPB	18 Apr 2019 16:00:02	S	Std
ICV08 - 1	ICV08	17572	3.9686	-7470A	PPB	18 Apr 2019 16:04:03	U	SMPL
ICB08 - 1	ICB08	367	0.0223	-7470A	PPB	18 Apr 2019 16:06:10	U	SMPL
CCV39 - 1	CCV39	21822	4.9434	-7470A	PPB	18 Apr 2019 16:08:16	U	SMPL
CCB39 - 1	CCB39	276	0.0015	-7470A	PPB	18 Apr 2019 16:10:21	U	SMPL
CRA - 1	CRA	1173	0.2072	-7470A	PPB	18 Apr 2019 16:12:28	U	SMPL
HighStd - 1	HighStd	43597	9.9378	-7470A	PPB	18 Apr 2019 16:14:33	U	SMPL
ChkStd - 1	ChkStd	31093	7.0698	-7470A	PPB	18 Apr 2019 16:16:40	U	SMPL
PB119017BL - 1	PBW	10	-0.0596	-7470A	PPB	18 Apr 2019 16:18:47	U	SMPL
PB119017BS - 1	LCSW	17236	3.8915	-7470A	PPB	18 Apr 2019 16:20:54	U	SMPL
K2196-06 - 1	HD-01-041619	116	-0.0352	-7470A	PPB	18 Apr 2019 16:23:01	U	SMPL
K2196-06DUP - 1	HD-01-041619DUP	179	-0.0208	-7470A	PPB	18 Apr 2019 16:25:09	U	SMPL
K2196-06MS - 1	HD-01-041619MS	16100	3.6309	-7470A	PPB	18 Apr 2019 16:27:16	U	SMPL
K2196-06MSD - 1	HD-01-041619MSD	16017	3.6119	-7470A	PPB	18 Apr 2019 16:29:24	U	SMPL
K2197-06 - 1	CL-01-041619-A	53	-0.0497	-7470A	PPB	18 Apr 2019 16:31:31	U	SMPL
CCV40 - 1	CCV40	22191	5.0280	-7470A	PPB	18 Apr 2019 16:33:36	U	SMPL
CCB40 - 1	CCB40	285	0.0035	-7470A	PPB	18 Apr 2019 16:35:43	U	SMPL
K2197-08 - 1	CL-01-041619-B	126	-0.0329	-7470A	PPB	18 Apr 2019 16:37:49	U	SMPL
K2197-10 - 1	CL-01-041619-C	127	-0.0327	-7470A	PPB	18 Apr 2019 16:39:56	U	SMPL
K2197-12 - 1	CL-01-041619-D	142	-0.0293	-7470A	PPB	18 Apr 2019 16:42:02	U	SMPL
K2197-14 - 1	CL-01-041619-E	220	-0.0114	-7470A	PPB	18 Apr 2019 16:44:09	U	SMPL
K2198-04 - 1	OK-02-041619	90	-0.0412	-7470A	PPB	18 Apr 2019 16:46:16	U	SMPL
K2198-06 - 1	OK-03-041719	384	0.0262	-7470A	PPB	18 Apr 2019 16:48:23	U	SMPL
K2430-02 - 1	OILY-STONES	95	-0.0401	-7470A	PPB	18 Apr 2019 16:50:31	U	SMPL
K2430-04 - 1	PAINT-CHIPS	91	-0.0410	-7470A	PPB	18 Apr 2019 16:52:39	U	SMPL
K2430-06 - 1	OILY-DEBRIS	94	-0.0403	-7470A	PPB	18 Apr 2019 16:54:47	U	SMPL
K2434-03 - 1	RO-288	112	-0.0362	-7470A	PPB	18 Apr 2019 16:56:53	U	SMPL
CCV41 - 1	CCV41	22046	4.9948	-7470A	PPB	18 Apr 2019 16:59:00	U	SMPL
CCB41 - 1	CCB41	325	0.0127	-7470A	PPB	18 Apr 2019 17:01:07	U	SMPL
K2435-03 - 1	BWM-S1-SOIL-PILE	112	-0.0362	-7470A	PPB	18 Apr 2019 17:03:14	U	SMPL
K2435-04 - 1	BWM-S2-SOIL-PILE	167	-0.0235	-7470A	PPB	18 Apr 2019 17:05:20	U	SMPL
K2441-04 - 1	TP-18-S	125	-0.0332	-7470A	PPB	18 Apr 2019 17:07:27	U	SMPL
K2442-04 - 1	TP-18-D	139	-0.0300	-7470A	PPB	18 Apr 2019 17:09:35	U	SMPL
K2442-08 - 1	TP-20-S	129	-0.0323	-7470A	PPB	18 Apr 2019 17:11:41	U	SMPL
K2442-12 - 1	TP-21-S	178	-0.0210	-7470A	PPB	18 Apr 2019 17:13:49	U	SMPL
K2442-16 - 1	TP-21-D	181	-0.0203	-7470A	PPB	18 Apr 2019 17:15:56	U	SMPL
PB119029BL - 1	PBW	340	0.0161	-7470A	PPB	18 Apr 2019 17:18:04	U	SMPL
PB119029BS - 1	LCSW	17428	3.9355	-7470A	PPB	18 Apr 2019 17:20:13	U	SMPL
K2241-09 - 1	MDL-MDL-WATER-03-QT2-2019	516	0.0565	-7470A	PPB	18 Apr 2019 17:22:19	U	SMPL
CCV42 - 1	CCV42	21343	4.8335	-7470A	PPB	18 Apr 2019 17:24:25	U	SMPL
CCB42 - 1	CCB42	347	0.0177	-7470A	PPB	18 Apr 2019 17:26:32	U	SMPL
PB119017TB - 1	PB118953TB	151	-0.0272	-7470A	PPB	18 Apr 2019 17:28:40	U	SMPL
K2196-06LX5 - 1		12	-0.0591	-7470A	PPB	18 Apr 2019 17:30:46	U	SMPL
K2196-06A - 1		16333	3.6844	-7470A	PPB	18 Apr 2019 17:32:53	U	SMPL
CCV43 - 1	CCV43	21869	4.9542	-7470A	PPB	18 Apr 2019 17:34:59	U	SMPL
CCB43 - 1	CCB43	191	-0.0180	-7470A	PPB	18 Apr 2019 17:37:07	U	SMPL