

LB102209


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61049



A= 0.0000e+000

B= 1.6494e-004

C= -1.6250e-002

Rho= 0.9999096

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	118	0.000	118				
0.2	0.200	0.178	-0.022	1177	0.0 %	1177				
2.5	2.500	2.558	0.058	15609	0.0 %	15609				
5.0	5.000	4.997	-0.003	30397	0.0 %	30397				
7.5	7.500	7.410	-0.090	45026	0.0 %	45026				
10.0	10.000	10.053	0.053	61049	0.0 %	61049				

LB102209 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Apr 2019 18:01:47

Sample ID	Extended ID	μ Abs	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50-0	118	-	0.00007471B	PPB	18 Apr 2019 18:02:36	S	Std
0.2 - 1	50-2	1177	-	0.20007471B	PPB	18 Apr 2019 18:04:53	S	Std
2.5 - 1	52-5	15609	-	2.50007471B	PPB	18 Apr 2019 18:07:11	S	Std
5.0 - 1	55-0	30397	-	5.00007471B	PPB	18 Apr 2019 18:09:28	S	Std
7.5 - 1	57-8	45026	-	7.50007471B	PPB	18 Apr 2019 18:11:45	S	Std
10.0 - 1	510-0	61049	-	10.00007471B	PPB	18 Apr 2019 18:14:01	S	Std
ICV09 - 1	ICV09	20841	3.4212	-7471B	PPB	18 Apr 2019 18:18:17	U	SMPL
ICB09 - 1	ICB09	0	-0.0162	-7471B	PPB	18 Apr 2019 18:20:33	U	SMPL
ICV09 - 1	ICV09	24496	4.0241	-7471B	PPB	18 Apr 2019 18:22:17	U	SMPL
ICB09 - 1	ICB09	31	-0.0111	-7471B	PPB	18 Apr 2019 18:24:32	U	SMPL
CCV44 - 1	CCV44	30381	4.9947	-7471B	PPB	18 Apr 2019 18:26:47	U	SMPL
CCB44 - 1	CCB44	39	-0.0098	-7471B	PPB	18 Apr 2019 18:29:03	U	SMPL
CRA - 1	CRA	1154	0.1741	-7471B	PPB	18 Apr 2019 18:31:19	U	SMPL
HighStd - 1	HighStd	60897	10.0280	-7471B	PPB	18 Apr 2019 18:33:34	U	SMPL
ChkStd - 1	ChkStd	41836	6.8841	-7471B	PPB	18 Apr 2019 18:35:50	U	SMPL
PB119013BL - 1	PBS	-7	-0.0174	-7471B	PPB	18 Apr 2019 18:38:07	U	SMPL
PB119013BS - 1	LCSS	22469	3.6897	-7471B	PPB	18 Apr 2019 18:40:23	U	SMPL
K1926-03 - 1	SUMP-MUCK	2331	0.3682	-7471B	PPB	18 Apr 2019 18:42:42	U	SMPL
K2193-07 - 1	HR-05-041719-A	3277	0.5243	-7471B	PPB	18 Apr 2019 18:44:59	U	SMPL
K2193-09 - 1	HR-05-041719-B	2590	0.4109	-7471B	PPB	18 Apr 2019 18:47:15	U	SMPL
K2193-11 - 1	HR-05-041719-C	4551	0.7344	-7471B	PPB	18 Apr 2019 18:49:32	U	SMPL
K2198-05 - 1	OK-03-041719	2817	0.4484	-7471B	PPB	18 Apr 2019 18:51:48	U	SMPL
CCV45 - 1	CCV45	30555	5.0234	-7471B	PPB	18 Apr 2019 18:54:03	U	SMPL
CCB45 - 1	CCB45	33	-0.0108	-7471B	PPB	18 Apr 2019 18:56:18	U	SMPL
K2444-01 - 1	FK-GF-02-037-A	5839	0.9468	-7471B	PPB	18 Apr 2019 18:58:34	U	SMPL
K2444-03 - 1	FK-GF-02-037-B	5919	0.9600	-7471B	PPB	18 Apr 2019 19:00:49	U	SMPL
K2453-01 - 1	OILY-STONES	1344	0.2054	-7471B	PPB	18 Apr 2019 19:03:05	U	SMPL
K2460-01 - 1	FK-EF-01-007-ABCD	2791	0.4441	-7471B	PPB	18 Apr 2019 19:05:21	U	SMPL
K2460-01DUP - 1	FK-EF-01-007-ABCDUP	2769	0.4405	-7471B	PPB	18 Apr 2019 19:07:37	U	SMPL
K2460-01MS - 1	FK-EF-01-007-ABCDMS	25802	4.2395	-7471B	PPB	18 Apr 2019 19:09:54	U	SMPL
K2460-01MSD - 1	FK-EF-01-007-ABCDMSD	25845	4.2466	-7471B	PPB	18 Apr 2019 19:12:10	U	SMPL
PB119014BL - 1	PBS	-8	-0.0176	-7471B	PPB	18 Apr 2019 19:14:28	U	SMPL
PB119014BS - 1	LCSS	22837	3.7504	-7471B	PPB	18 Apr 2019 19:16:45	U	SMPL
K2241-03 - 1	MDL-MDL-SOIL-03-QT2-2019	408	0.0510	-7471B	PPB	18 Apr 2019 19:23:02	U	SMPL
CCV46 - 1	CCV46	30758	5.0569	-7471B	PPB	18 Apr 2019 19:25:18	U	SMPL
CCB46 - 1	CCB46	106	0.0012	-7471B	PPB	18 Apr 2019 19:27:33	U	SMPL
K2460-01LX5 - 1		756	0.1084	-7471B	PPB	18 Apr 2019 19:29:49	U	SMPL
K2460-01A - 1		25867	4.2502	-7471B	PPB	18 Apr 2019 19:32:05	U	SMPL
CCV47 - 1	CCV47	30894	5.0793	-7471B	PPB	18 Apr 2019 19:34:21	U	SMPL
CCB47 - 1	CCB47	-12	-0.0182	-7471B	PPB	18 Apr 2019 19:36:36	U	SMPL