



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102219

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/19/2019	10:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/19/2019	10:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/19/2019	10:18
4	ICV	1	Water	NA	NA	20.1	6.99	04/19/2019	10:19
5	CCV1	1	Water	NA	NA	20.1	2.01	04/19/2019	10:20
6	K2487-01	1	Water	NA	NA	14.2	6.49	04/19/2019	10:21
7	K2487-02	1	Water	NA	NA	12.9	6.92	04/19/2019	10:23
8	K2487-02DUP	1	Water	NA	NA	12.8	6.93	04/19/2019	10:25
9	K2487-05	1	Water	NA	NA	13.4	6.97	04/19/2019	10:26
10	K2487-06	1	Water	NA	NA	16.2	5.62	04/19/2019	10:27
11	K2494-01	1	Water	NA	NA	21.0	5.02	04/19/2019	10:30
12	CCV2	1	Water	NA	NA	21.0	12.01	04/19/2019	10:31

WORKLIST(Hardcopy Internal Chain)

LB 102219

WorkList Name : PH K2487

WorkList ID : 125066

Date : 04-19-2019 10:00:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/18/2019	Water	K2487-01-1A	pH	Cool 4 deg C	WOOD06	K51	GMW-10	04/18/2019	SM 4500-H B
04/18/2019	Water	K2487-02 1A	pH	Cool 4 deg C	WOOD06	K51	GMW-7	04/18/2019	SM 4500-H B
04/18/2019	Water	K2487-05 1A	pH	Cool 4 deg C	WOOD06	K51	GMW-X	04/18/2019	SM 4500-H B
04/18/2019	Water	K2487-06 1A	pH	Cool 4 deg C	WOOD06	K51	FB	04/18/2019	SM 4500-H B
04/19/2019	Water	K2494-01 1A	pH	Cool 4 deg C	STER10	K32	CELL-A-POOL-WATER(DE-IC	04/19/2019	SM 4500-H B

Date/Time 04.19.19 10:05
 Received by: SH
 Relinquished by: SH

Date/Time 04.19.19 10:20
 Received by: SH
 Relinquished by: SH