



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/19/2019

Run Number: LB102224

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 15028095

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 04/18/2019 16:00 TEMP OUT: 104 04/19/2019 07:00
TEMP2 IN: 104 04/19/2019 07:45 TEMP OUT: 104 04/19/2019 09:00
TEMP3 IN: 105 04/19/2019 10:00 TEMP OUT: 105 04/19/2019 11:40
TEMP4 IN: 105 04/19/2019 12:30 TEMP OUT: 105 04/19/2019 14:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB102224BL	LB102224BL	1.4025	1.4025	100	1.4025	1.4025	1.4025	0.0000	0
2	LB102224BS	LB102224BS	1.3965	1.3965	100	1.4470	1.4470	1.4470	0.0505	505
3	K2467-01	SPDES-1	1.4052	1.4052	900	1.4058	1.4058	1.4058	0.0006	0.7
4	K2479-01	TP-TRACK-8-9	1.4077	1.4077	300	1.4130	1.4130	1.4130	0.0053	17.7
5	K2479-01DUP	TP-TRACK-8-9DUP	1.4088	1.4088	300	1.4140	1.4140	1.4140	0.0052	17.3
6	K2479-02	TP-TRACK-70-71	1.4000	1.4000	900	1.4032	1.4032	1.4032	0.0032	3.6
7	K2479-03	TP-TRACK-74-74	1.4082	1.4082	900	1.4097	1.4097	1.4097	0.0015	1.7
8	K2479-04	TP-TRACK-78-79	1.4082	1.4082	900	1.4098	1.4098	1.4098	0.0016	1.8
9	K2482-01	AU-041819	1.3916	1.3916	10	1.4719	1.4719	1.4719	0.0803	8030

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS041919

WorkList ID : 125031

Date : 04-19-2019 07:01:40

2/13/02224

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/19/2019	Water	K2467-01	TSS	Cool 4 deg C	SKAN06	K23	SPDES-1	04/18/2019	SM2540 D
04/25/2019	Water	K2479-01	TSS	Cool 4 deg C	TULL02	K41	TP-TRACK-8-9	04/18/2019	SM2540 D
04/25/2019	Water	K2479-02	TSS	Cool 4 deg C	TULL02	K41	TP-TRACK-70-71	04/18/2019	SM2540 D
04/25/2019	Water	K2479-03	TSS	Cool 4 deg C	TULL02	K41	TP-TRACK-74-74	04/18/2019	SM2540 D
04/25/2019	Water	K2479-04	TSS	Cool 4 deg C	TULL02	K41	TP-TRACK-78-79	04/18/2019	SM2540 D
04/25/2019	Water	K2482-01	TSS	Cool 4 deg C	PSEG01	K11	AU-041819	04/18/2019	SM2540 D

Date/Time 04/19/19 07:20
 Received by: 
 Relinquished by: _____

Date/Time 04/19/19 10:40
 Received by: 
 Relinquished by: _____