

LB102225


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

63135



A= 0.0000e+000

B= 1.5938e-004

C= -7.0069e-002

Rho= 0.9990366

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.084	-0.084	-86	0.000	-86				
0.2	0.200	0.215	0.015	1789	0.0 %	1789				
2.5	2.500	2.471	-0.029	15944	0.0 %	15944				
5.0	5.000	5.321	0.321	33822	0.0 %	33822				
7.5	7.500	7.284	-0.216	46144	0.0 %	46144				
10.0	10.000	9.993	-0.007	63135	0.0 %	63135				

LB102225 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Apr 2019 11:35:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	500	-86	-	0.00007471B	PPB		19 Apr 2019 12:02:23	S	Std
0.2 - 1	502	1789	-	0.20007471B	PPB		19 Apr 2019 12:04:40	S	Std
2.5 - 1	525	15944	-	2.50007471B	PPB		19 Apr 2019 12:06:57	S	Std
5.0 - 1	550	33822	-	5.00007471B	PPB		19 Apr 2019 12:09:15	S	Std
7.5 - 1	575	46144	-	7.50007471B	PPB		19 Apr 2019 12:11:33	S	Std
10.0 - 1	5100	63135	-	10.00007471B	PPB		19 Apr 2019 12:13:49	S	Std
ICV10 - 1	ICV10	24622	3.8542	-7471B	PPB		19 Apr 2019 12:17:04	U	SMPL
ICB10 - 1	ICB10	-71	-0.0814	-7471B	PPB		19 Apr 2019 12:19:19	U	SMPL
CCV48 - 1	CCV48	32722	5.1452	-7471B	PPB		19 Apr 2019 12:21:35	U	SMPL
CCB48 - 1	CCB48	-50	-0.0780	-7471B	PPB		19 Apr 2019 12:23:50	U	SMPL
CRA - 1	CRA	1804	0.2175	-7471B	PPB		19 Apr 2019 12:26:06	U	SMPL
HighStd - 1	HighStd	63883	10.1118	-7471B	PPB		19 Apr 2019 12:28:21	U	SMPL
ChkStd - 1	ChkStd	43244	6.8223	-7471B	PPB		19 Apr 2019 12:30:36	U	SMPL
PB119062BL - 1	PBS	-22	-0.0736	-7471B	PPB		19 Apr 2019 12:32:53	U	SMPL
PB119062BS - 1	LCSS	22177	3.4646	-7471B	PPB		19 Apr 2019 12:35:10	U	SMPL
K2199-13 - 1	OR-03-041819-A	3876	0.5477	-7471B	PPB		19 Apr 2019 12:37:27	U	SMPL
K2199-13DUP - 1	OR-03-041819-ADUP	3975	0.5635	-7471B	PPB		19 Apr 2019 12:39:44	U	SMPL
K2199-13MS - 1	OR-03-041819-AMS	27722	4.3483	-7471B	PPB		19 Apr 2019 12:42:01	U	SMPL
K2199-13MSD - 1	OR-03-041819-AMSD	27836	4.3665	-7471B	PPB		19 Apr 2019 12:44:18	U	SMPL
K2199-15 - 1	OR-03-041819-B	2587	0.3423	-7471B	PPB		19 Apr 2019 12:46:34	U	SMPL
CCV49 - 1	CCV49	33051	5.1977	-7471B	PPB		19 Apr 2019 12:48:49	U	SMPL
CCB49 - 1	CCB49	15	-0.0677	-7471B	PPB		19 Apr 2019 12:51:04	U	SMPL
K2201-01 - 1	PL-01-041819-A	1232	0.1263	-7471B	PPB		19 Apr 2019 12:53:20	U	SMPL
K2201-03 - 1	PL-01-041819-B	3990	0.5659	-7471B	PPB		19 Apr 2019 12:55:36	U	SMPL
K2201-05 - 1	PL-01-041819-C	7926	1.1932	-7471B	PPB		19 Apr 2019 12:57:52	U	SMPL
K2201-07 - 1	PL-01-041819-D	5472	0.8021	-7471B	PPB		19 Apr 2019 13:00:07	U	SMPL
K2201-09 - 1	PL-01-041819-E	1049	0.0971	-7471B	PPB		19 Apr 2019 13:02:23	U	SMPL
K2202-01 - 1	SA-06-041819-A	2185	0.2782	-7471B	PPB		19 Apr 2019 13:04:39	U	SMPL
K2202-03 - 1	SA-06-041819-B	2735	0.3658	-7471B	PPB		19 Apr 2019 13:06:56	U	SMPL
K2203-13 - 1	SU-02-041819-A	1543	0.1759	-7471B	PPB		19 Apr 2019 13:09:14	U	SMPL
K2203-15 - 1	SU-02-041819-B	1772	0.2124	-7471B	PPB		19 Apr 2019 13:11:31	U	SMPL
K2472-01 - 1	EXCAVATION-NEWARK-SOUTH-PILE	53418	8.4438	-7471B	PPB		19 Apr 2019 13:13:47	U	SMPL
CCV50 - 1	CCV50	33845	5.3242	-7471B	PPB		19 Apr 2019 13:16:02	U	SMPL
CCB50 - 1	CCB50	109	-0.0527	-7471B	PPB		19 Apr 2019 13:18:19	U	SMPL
K2475-01 - 1	TP-TRACK-8-9	3450	0.4798	-7471B	PPB		19 Apr 2019 13:20:36	U	SMPL
K2475-03 - 1	TP-TRACK-70-71	2362	0.3064	-7471B	PPB		19 Apr 2019 13:22:51	U	SMPL
K2475-05 - 1	TP-TRACK-74-75	4333	0.6205	-7471B	PPB		19 Apr 2019 13:25:08	U	SMPL
K2475-07 - 1	TP-TRACK-78-79	2749	0.3681	-7471B	PPB		19 Apr 2019 13:27:24	U	SMPL
K2484-01 - 1	72-12000	878	0.0699	-7471B	PPB		19 Apr 2019 13:29:39	U	SMPL
K2484-03 - 1	200027	1210	0.1228	-7471B	PPB		19 Apr 2019 13:31:56	U	SMPL
K2484-05 - 1	RT-1873	1135	0.1108	-7471B	PPB		19 Apr 2019 13:34:12	U	SMPL
K2485-01 - 1	SOIL-CUTTINGS	1001	0.0895	-7471B	PPB		19 Apr 2019 13:36:29	U	SMPL
K2199-13LX5 - 1		946	0.0807	-7471B	PPB		19 Apr 2019 13:38:47	U	SMPL
K2199-13A - 1		28493	4.4712	-7471B	PPB		19 Apr 2019 13:41:03	U	SMPL
CCV51 - 1	CCV51	33618	5.2881	-7471B	PPB		19 Apr 2019 13:43:19	U	SMPL
CCB51 - 1	CCB51	42	-0.0634	-7471B	PPB		19 Apr 2019 13:45:35	U	SMPL