



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102241

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/19/2019	09:20
2	K2482-02		1	9	100.00	>212.0	>212.0	04/19/2019	09:45
3	K2482-02DUP		1	9	100.00	>212.0	>212.0	04/19/2019	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP41919

WorkList ID : 125061

Date : 04-19-2019 08:45:10

LB102241

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2019	Water	K2482-02	Flash Point	Cool 4 deg C	PSEG01	K11	AU-041819	04/18/2019	1010A

Date/Time 04/19/19 08:50
 Received by: KS
 Relinquished by: CP

Date/Time 04/19/19 10:30
 Received by: CP
 Relinquished by: KS