



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102253

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/19/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	04/19/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.02	04/19/2019	16:02
4	ICV	1	Water	NA	NA	20.1	7.00	04/19/2019	16:03
5	CCV1	1	Water	NA	NA	20.1	2.01	04/19/2019	16:05
6	K2482-02	1	Water	NA	NA	22.1	6.31	04/19/2019	16:06
7	K2516-02	1	Water	NA	NA	21.6	7.74	04/19/2019	16:08
8	K2516-02DUP	1	Water	NA	NA	21.7	7.75	04/19/2019	16:10
9	CCV2	1	Water	NA	NA	20.1	12.02	04/19/2019	16:11

18102253

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K2582

WorkList ID : 125099

Date : 04-19-2019 14:58:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/18/2019	Water	K2482-02	✓✗ Corrosivity	Cool 4 deg C	PSEG01	K11	AU-041819	04/18/2019	9040C
04/19/2019	Water	K2516-02	✗ Corrosivity	Cool 4 deg C	PSEG01	K11	P041919	04/19/2019	9040C

Date/Time 04-19-19 15:45
 Received by: Jo
 Relinquished by: ed

Date/Time 04-19-19 16:20
 Received by: ew
 Relinquished by: Jg