


LB102263

7470A

INSTRUMENT ID: CUI

Linear



A= 0.0000e+000

B= 2.3433e-004

C= -6.3475e-002

Rho= 0.9998895

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	219	0.000	219				
0.2	0.200	0.187	-0.013	1070	0.0 %	1070				
2.5	2.500	2.461	-0.039	10772	0.0 %	10772				
5.0	5.000	5.120	0.120	22120	0.0 %	22120				
7.5	7.500	7.484	-0.016	32208	0.0 %	32208				
10.0	10.000	9.960	-0.040	42777	0.0 %	42777				

LB102263 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Apr 2019 15:38:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	219	-	0.0000	7470A	PPB	22 Apr 2019 15:39:15	S	Std
0.2 - 1	50.2	1070	-	0.2000	7470A	PPB	22 Apr 2019 15:41:23	S	Std
2.5 - 1	52.5	10772	-	2.5000	7470A	PPB	22 Apr 2019 15:43:31	S	Std
5.0 - 1	55.0	22120	-	5.0000	7470A	PPB	22 Apr 2019 15:45:39	S	Std
7.5 - 1	57.5	32208	-	7.5000	7470A	PPB	22 Apr 2019 15:47:47	S	Std
10.0 - 1	510.0	42777	-	10.0000	7470A	PPB	22 Apr 2019 15:49:54	S	Std
ICV13 - 1	ICV13	17174	3.9609	-	7470A	PPB	22 Apr 2019 15:52:33	U	SMPL
ICB13 - 1	ICB13	148	-0.0288	-	7470A	PPB	22 Apr 2019 15:54:39	U	SMPL
CCV65 - 1	CCV65	22177	5.1333	-	7470A	PPB	22 Apr 2019 15:56:45	U	SMPL
CCB65 - 1	CCB65	193	-0.0182	-	7470A	PPB	22 Apr 2019 15:58:51	U	SMPL
CRA - 1	CRA	1089	0.1917	-	7470A	PPB	22 Apr 2019 16:00:57	U	SMPL
HighStd - 1	HighStd	43649	10.1648	-	7470A	PPB	22 Apr 2019 16:03:03	U	SMPL
ChkStd - 1	ChkStd	30298	7.0363	-	7470A	PPB	22 Apr 2019 16:05:09	U	SMPL
PB119119BL - 1	PBW	108	-0.0382	-	7470A	PPB	22 Apr 2019 16:07:16	U	SMPL
PB119119BS - 1	LCSW	16933	3.9044	-	7470A	PPB	22 Apr 2019 16:09:23	U	SMPL
K2482-02 - 1	AU-041819	195	-0.0178	-	7470A	PPB	22 Apr 2019 16:11:30	U	SMPL
K2516-02 - 1	P041919	195	-0.0178	-	7470A	PPB	22 Apr 2019 16:13:38	U	SMPL
K2516-02DUP - 1	P041919DUP	184	-0.0204	-	7470A	PPB	22 Apr 2019 16:15:46	U	SMPL
K2516-02MS - 1	P041919MS	17787	4.1045	-	7470A	PPB	22 Apr 2019 16:17:53	U	SMPL
K2516-02MSD - 1	P041919MSD	17739	4.0933	-	7470A	PPB	22 Apr 2019 16:19:59	U	SMPL
CCV66 - 1	CCV66	22224	5.1443	-	7470A	PPB	22 Apr 2019 16:22:05	U	SMPL
CCB66 - 1	CCB66	168	-0.0241	-	7470A	PPB	22 Apr 2019 16:24:11	U	SMPL
PB119120BL - 1	PBW	173	-0.0229	-	7470A	PPB	22 Apr 2019 16:26:17	U	SMPL
PB119120BS - 1	LCSW	16785	3.8698	-	7470A	PPB	22 Apr 2019 16:28:24	U	SMPL
K2463-01 - 1	DRUM-1	87	-0.0431	-	7470A	PPB	22 Apr 2019 16:30:30	U	SMPL
K2463-02 - 1	DRUM-2	125	-0.0342	-	7470A	PPB	22 Apr 2019 16:32:36	U	SMPL
K2464-02 - 1	M-COMP-4.18.19	136	-0.0316	-	7470A	PPB	22 Apr 2019 16:34:43	U	SMPL
K2465-02 - 1	Q-COMP-4.18.19	143	-0.0300	-	7470A	PPB	22 Apr 2019 16:36:50	U	SMPL
K2501-01 - 1	T-2	193	-0.0182	-	7470A	PPB	22 Apr 2019 16:38:57	U	SMPL
K2508-01 - 1	OILY-DEBRIS-220-6	175	-0.0225	-	7470A	PPB	22 Apr 2019 16:41:05	U	SMPL
K2508-03 - 1	OIL-220-6	448	0.0415	-	7470A	PPB	22 Apr 2019 16:43:13	U	SMPL
K2508-08 - 1	OILY-DEBRIS-220-3	110	-0.0377	-	7470A	PPB	22 Apr 2019 16:45:20	U	SMPL
CCV67 - 1	CCV67	22060	5.1058	-	7470A	PPB	22 Apr 2019 16:47:26	U	SMPL
CCB67 - 1	CCB67	108	-0.0382	-	7470A	PPB	22 Apr 2019 16:49:32	U	SMPL
K2508-10 - 1	OIL-220-3	329	0.0136	-	7470A	PPB	22 Apr 2019 16:51:39	U	SMPL
K2510-02 - 1	SOIL-PILE-A	91	-0.0422	-	7470A	PPB	22 Apr 2019 16:53:45	U	SMPL
K2510-04 - 1	SOIL-PILE-B	165	-0.0248	-	7470A	PPB	22 Apr 2019 16:55:53	U	SMPL
K2515-04 - 1	TP-15	149	-0.0286	-	7470A	PPB	22 Apr 2019 16:58:00	U	SMPL
K2515-08 - 1	TP-10	164	-0.0250	-	7470A	PPB	22 Apr 2019 17:00:06	U	SMPL
K2515-12 - 1	TP-9	169	-0.0239	-	7470A	PPB	22 Apr 2019 17:02:14	U	SMPL
K2515-16 - 1	TP-12	149	-0.0286	-	7470A	PPB	22 Apr 2019 17:04:21	U	SMPL
K2515-20 - 1	TP-13	220	-0.0119	-	7470A	PPB	22 Apr 2019 17:06:29	U	SMPL
K2515-24 - 1	TP-16	228	-0.0100	-	7470A	PPB	22 Apr 2019 17:08:37	U	SMPL
K2515-24DUP - 1	TP-16DUP	213	-0.0136	-	7470A	PPB	22 Apr 2019 17:10:43	U	SMPL
CCV68 - 1	CCV68	21433	4.9589	-	7470A	PPB	22 Apr 2019 17:12:50	U	SMPL
CCB68 - 1	CCB68	168	-0.0241	-	7470A	PPB	22 Apr 2019 17:14:57	U	SMPL
K2515-24MS - 1	TP-16MS	17232	3.9745	-	7470A	PPB	22 Apr 2019 17:17:03	U	SMPL
K2515-24MSD - 1	TP-16MSD	17339	3.9996	-	7470A	PPB	22 Apr 2019 17:19:10	U	SMPL
PB119121BL - 1	PBW	104	-0.0391	-	7470A	PPB	22 Apr 2019 17:21:16	U	SMPL
PB119121BS - 1	LCSW	15867	3.6546	-	7470A	PPB	22 Apr 2019 17:23:22	U	SMPL
K2487-01 - 1	GMW-10	8106	1.8360	-	7470A	PPB	22 Apr 2019 17:25:29	U	SMPL
K2487-02 - 1	GMW-7	7260	1.6378	-	7470A	PPB	22 Apr 2019 17:27:36	U	SMPL
K2487-02DUP - 1	GMW-7DUP	7244	1.6340	-	7470A	PPB	22 Apr 2019 17:29:44	U	SMPL
K2487-03 - 1	K2487-02MS	24267	5.6230	-	7470A	PPB	22 Apr 2019 17:31:52	U	SMPL
K2487-04 - 1	K2487-02MSD	24814	5.7512	-	7470A	PPB	22 Apr 2019 17:34:01	U	SMPL
K2487-05 - 1	GMW-X	3523	0.7621	-	7470A	PPB	22 Apr 2019 17:36:07	U	SMPL
CCV69 - 1	CCV69	21829	5.0517	-	7470A	PPB	22 Apr 2019 17:38:14	U	SMPL
CCB69 - 1	CCB69	155	-0.0272	-	7470A	PPB	22 Apr 2019 17:40:20	U	SMPL
K2487-06 - 1	FB	155	-0.0272	-	7470A	PPB	22 Apr 2019 17:42:26	U	SMPL
K2487-08 - 1	GMW-10	191	-0.0187	-	7470A	PPB	22 Apr 2019 17:44:32	U	SMPL
K2487-09 - 1	GMW-7	173	-0.0229	-	7470A	PPB	22 Apr 2019 17:46:39	U	SMPL
K2487-10 - 1	K2487-09MS	16129	3.7160	-	7470A	PPB	22 Apr 2019 17:48:47	U	SMPL
K2487-11 - 1	K2487-09MSD	15950	3.6741	-	7470A	PPB	22 Apr 2019 17:50:55	U	SMPL
K2487-12 - 1	GMW-X	171	-0.0234	-	7470A	PPB	22 Apr 2019 17:53:02	U	SMPL
K2487-13 - 1	FB	27	-0.0571	-	7470A	PPB	22 Apr 2019 17:55:09	U	SMPL
K2516-01 - 1	P041919	357	0.0202	-	7470A	PPB	22 Apr 2019 17:57:17	U	SMPL
PB119119TB - 1	PB119057TB	189	-0.0192	-	7470A	PPB	22 Apr 2019 17:59:26	U	SMPL
PB119120TB - 1	PB119037TB	169	-0.0239	-	7470A	PPB	22 Apr 2019 18:01:34	U	SMPL
CCV70 - 1	CCV70	21337	4.9364	-	7470A	PPB	22 Apr 2019 18:03:41	U	SMPL
CCB70 - 1	CCB70	165	-0.0248	-	7470A	PPB	22 Apr 2019 18:05:48	U	SMPL
K2516-02LX5 - 1		-13	-0.0665	-	7470A	PPB	22 Apr 2019 18:07:55	U	SMPL
K2516-02A - 1		16751	3.8618	-	7470A	PPB	22 Apr 2019 18:10:02	U	SMPL
K2515-24LX5 - 1		-2	-0.0639	-	7470A	PPB	22 Apr 2019 18:12:09	U	SMPL
K2515-24A - 1		17341	4.0000	-	7470A	PPB	22 Apr 2019 18:14:16	U	SMPL
K2487-02LX5 - 1		1185	0.2142	-	7470A	PPB	22 Apr 2019 18:16:23	U	SMPL
K2487-02A - 1		24755	5.7374	-	7470A	PPB	22 Apr 2019 18:18:30	U	SMPL
K2487-09A - 1		16217	3.7367	-	7470A	PPB	22 Apr 2019 18:20:38	U	SMPL
CCV71 - 1	CCV71	20840	4.8200	-	7470A	PPB	22 Apr 2019 18:22:46	U	SMPL
CCB71 - 1	CCB71	159	-0.0262	-	7470A	PPB	22 Apr 2019 18:24:54	U	SMPL