



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102266

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/22/2019	11:40
2	K2516-02		1	9	100.00	>212.0	>212.0	04/22/2019	12:05
3	K2516-02DUP		1	9	100.00	>212.0	>212.0	04/22/2019	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102266

WorkList Name : FP42219

WorkList ID : 125143

Date : 04-22-2019 11:00:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2019	Water	K2516-02	Flash Point	Cool 4 deg C	PSEG01	K11	P041919	04/19/2019	1010A

Date/Time 04/22/19 11:10
 Received by: KS
 Relinquished by: KS

Date/Time 04/22/19 12:40
 Received by: KS
 Relinquished by: KS