



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102272

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/22/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	7.01	04/22/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.04	04/22/2019	16:32
4	ICV	1	Water	NA	NA	20.1	6.99	04/22/2019	16:33
5	CCV1	1	Water	NA	NA	20.2	2.01	04/22/2019	16:34
6	K2522-01	1	Water	NA	NA	18.8	6.97	04/22/2019	16:35
7	K2522-01DUP	1	Water	NA	NA	18.9	6.98	04/22/2019	16:36
8	K2522-04	1	Water	NA	NA	19.5	6.09	04/22/2019	16:37
9	CCV2	1	Water	NA	NA	20.2	12.01	04/22/2019	16:39

191022#2

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2522 W

WorkList ID : 125156

Date : 04-22-2019 16:01:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/22/2019	Water	K2522-01	pH	Cool 4 deg C	PSEG01	K12	WELL-1	04/22/2019	9040C
04/22/2019	Water	K2522-04	pH	Cool 4 deg C	PSEG01	K12	WELL-2	04/22/2019	9040C

Date/Time 04.22.19 16:10
Received by: Jn
Relinquished by: CD

Date/Time 04.22.19 17:10
Received by: JP
Relinquished by: CD