



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromiu
Run Number: LB102278

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards like HexaChrome 1 PPM, Hexa Chrome 0.5 PPM, hexchrome 0.1 ppm, etc.

Intercept: 0.0004 Slope: 0.7717 Regression: 0.999986

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB102278

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH		Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
						HN03	H2SO4	Backgrnd	Color				
1	ICV	0.5	1	100	100		1.87	0.000	0.392	0.392	0.507	04/22/2019	14:03
2	ICB		1	100	100		1.62	0.000	0.000	0.000	-0.001	04/22/2019	14:03
3	CCV1	0.5	1	100	100		2.15	0.000	0.393	0.393	0.509	04/22/2019	14:04
4	CCB1		1	100	100		2.39	0.000	0.000	0.000	-0.001	04/22/2019	14:04
5	PB119118BL		1	2.55	100	7.39	1.74	0.000	0.000	0.000	-0.001	04/22/2019	14:05
6	PB119118BS	20	1	2.50	100	7.28	1.53	0.000	0.369	0.369	0.478	04/22/2019	14:05
7	K2241-03		1	2.51	100	7.12	1.77	0.000	0.005	0.005	0.006	04/22/2019	14:06
8	K2472-01		1	2.50	100	7.03	2.24	0.016	0.016	0.000	-0.001	04/22/2019	14:06
9	K2484-01		1	2.55	100	7.19	2.39	0.189	0.189	0.000	-0.001	04/22/2019	14:07
10	K2484-03		1	2.53	100	7.24	1.73	0.013	0.016	0.003	0.003	04/22/2019	14:07
11	K2484-05		1	2.50	100	7.36	2.08	0.011	0.011	0.000	-0.001	04/22/2019	14:08
12	K2510-01		1	2.54	100	7.20	2.41	0.016	0.016	0.000	-0.001	04/22/2019	14:08
13	K2510-03		1	2.53	100	7.12	2.16	0.013	0.013	0.000	-0.001	04/22/2019	14:09
14	K2514-01		1	2.51	100	7.96	2.25	0.018	0.018	0.000	-0.001	04/22/2019	14:09
15	CCV2	0.5	1	100	100		1.87	0.000	0.393	0.393	0.509	04/22/2019	14:10
16	CCB2		1	100	100		1.92	0.000	0.000	0.000	-0.001	04/22/2019	14:10
17	K2515-03		1	2.56	100	7.41	2.04	0.020	0.020	0.000	-0.001	04/22/2019	14:11
18	K2515-07		1	2.52	100	7.58	2.37	0.015	0.015	0.000	-0.001	04/22/2019	14:11
19	K2515-11		1	2.54	100	7.16	1.58	0.020	0.020	0.000	-0.001	04/22/2019	14:12
20	K2515-15		1	2.53	100	7.02	1.63	0.011	0.011	0.000	-0.001	04/22/2019	14:12
21	K2515-15DUP		1	2.50	100	7.13	1.95	0.010	0.010	0.000	-0.001	04/22/2019	14:13
22	K2515-15MSPre	40	2	2.51	100	7.69	2.07	0.012	0.355	0.343	0.444	04/22/2019	14:13
23	K2515-15MS2Ins	1284	40	2.52	100	7.47	2.38	0.011	0.592	0.581	0.752	04/22/2019	14:14
24	K2515-15MS3Post	40	2	2.51	100	7.76	1.64	0.014	0.360	0.346	0.448	04/22/2019	14:14
25	K2515-19		1	2.50	100	7.32	1.82	0.023	0.023	0.000	-0.001	04/22/2019	14:15
26	K2515-23		1	2.53	100	7.06	1.74	0.032	0.032	0.000	-0.001	04/22/2019	14:15
27	CCV3	0.5	1	100	100		2.03	0.000	0.393	0.393	0.509	04/22/2019	14:16
28	CCB3		1	100	100		2.21	0.000	0.000	0.000	-0.001	04/22/2019	14:16