

LB102292


245.1

INSTRUMENT ID: CU1

Linear

 μ Abs.:

44362



A= 0.0000e+000

B= 2.2636e-004

C= -1.9114e-002

Rho= 0.9999751

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	108	0.000	108				
0.05	0.050	0.051	0.001	308	0.0 %	308				
0.2	0.200	0.224	0.024	1072	0.0 %	1072				
2.5	2.500	2.456	-0.044	10936	0.0 %	10936				
5.0	5.000	5.025	0.025	22285	0.0 %	22285				
7.5	7.500	7.466	-0.034	33067	0.0 %	33067				
10.0	10.000	10.023	0.023	44362	0.0 %	44362				

LB102292 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 Apr 2019 11:16:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	108	-	0.0000	245.1	PPB	23 Apr 2019 11:16:57	S	Std
0.05 - 1	50.05	308	-	0.0500	245.1	PPB	23 Apr 2019 11:19:14	S	Std
0.2 - 1	50.2	1072	-	0.2000	245.1	PPB	23 Apr 2019 11:21:32	S	Std
2.5 - 1	52.5	10936	-	2.5000	245.1	PPB	23 Apr 2019 11:23:50	S	Std
5.0 - 1	55.0	22285	-	5.0000	245.1	PPB	23 Apr 2019 11:26:07	S	Std
7.5 - 1	57.5	33067	-	7.5000	245.1	PPB	23 Apr 2019 11:28:23	S	Std
10.0 - 1	510.0	44362	-	10.0000	245.1	PPB	23 Apr 2019 11:30:39	S	Std
ICV16 - 1	ICV16	17312	3.8997	-245.1		PPB	23 Apr 2019 11:34:37	U	SMPL
ICB16 - 1	ICB16	8	-0.0173	-245.1		PPB	23 Apr 2019 11:36:52	U	SMPL
CCV81 - 1	CCV81	22382	5.0473	-245.1		PPB	23 Apr 2019 11:39:07	U	SMPL
CCB81 - 1	CCB81	30	-0.0123	-245.1		PPB	23 Apr 2019 11:41:22	U	SMPL
CRA - 1	CRA	290	0.0465	-245.1		PPB	23 Apr 2019 11:43:37	U	SMPL
HighStd - 1	HighStd	43901	9.9184	-245.1		PPB	23 Apr 2019 11:45:52	U	SMPL
ChkStd - 1	ChkStd	31017	7.0020	-245.1		PPB	23 Apr 2019 11:48:08	U	SMPL
PB119143BL - 1	PBW	36	-0.0110	-245.1		PPB	23 Apr 2019 11:50:24	U	SMPL
PB119143BS - 1	LCSW	17561	3.9560	-245.1		PPB	23 Apr 2019 11:52:40	U	SMPL
K2241-12 - 1	MDL-MDL-WATER-06-QT2-2019	289	0.0463	-245.1		PPB	23 Apr 2019 11:58:19	U	SMPL
CCV82 - 1	CCV82	22321	5.0335	-245.1		PPB	23 Apr 2019 12:04:19	U	SMPL
CCB82 - 1	CCB82	-7	-0.0207	-245.1		PPB	23 Apr 2019 12:06:37	U	SMPL