



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102296

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/23/2019	09:45
2	K2522-02		1	9	100.00	>212.0	>212.0	04/23/2019	10:20
3	K2522-02DUP		1	9	100.00	>212.0	>212.0	04/23/2019	10:50
4	K2522-05		1	7	100.00	>212.0	>212.0	04/23/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102296

WorkList Name : FP42319

WorkList ID : 125197

Date : 04-23-2019 09:16:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2019	Water	K2522-02	C Flash Point	Cool 4 deg C	PSEG01	K12	WELL-1	04/22/2019	1010A
04/26/2019	Water	K2522-05	C Flash Point	Cool 4 deg C	PSEG01	K12	WELL-2	04/22/2019	1010A

Date/Time 04/23/19 09:20
 Received by: KS
 Relinquished by: CS

Date/Time 04/23/19 11:25
 Received by: CS
 Relinquished by: KS