

LB102297


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52681



A= 0.0000e+000

B= 1.9220e-004

C= -5.2129e-002

Rho= 0.9998480

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.001	-0.001	265	0.000	265					—
0.05	0.050										
0.20	0.200	0.214	0.014	1383	0.0 %	1383					7
2.50	2.500	2.574	0.074	13666	0.0 %	13666					3
5.00	5.000	4.898	-0.102	25754	0.0 %	25754					-2
7.50	7.500	7.442	-0.058	38994	0.0 %	38994					-1
10.0	10.000	10.073	0.073	52681	0.0 %	52681					1

LB 04-23-19

LB102297

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	265	-		0 ISM02.4	PPB	4/23/2019 13:52	Std
	0.2 S0.2	1383	-		0.2 ISM02.4	PPB	4/23/2019 13:55	Std
	2.5 S2.5	13666	-		2.5 ISM02.4	PPB	4/23/2019 13:57	Std
	5 S5.0	25754	-		5 ISM02.4	PPB	4/23/2019 13:59	Std
	7.5 S7.5	38994	-		7.5 ISM02.4	PPB	4/23/2019 14:02	Std
	10 S10.0	52681	-		10 ISM02.4	PPB	4/23/2019 14:04	Std
ICV	ICV	20501	3.8881 -		ISM02.4	PPB	4/23/2019 14:08	SMPL
ICB	ICB	164	-0.0206 -		ISM02.4	PPB	4/23/2019 14:10	SMPL
CCV083	CCV083	25730	4.8931 -		ISM02.4	PPB	4/23/2019 14:12	SMPL
CCB083	CCB083	169	-0.0196 -		ISM02.4	PPB	4/23/2019 14:15	SMPL
PB119142BL	PBS142	212	-0.0114 -		ISM02.4	PPB	4/23/2019 14:17	SMPL
K2474-01	MA5125	1258	0.1897 -		ISM02.4	PPB	4/23/2019 14:19	SMPL
K2474-02	MA5131	317	0.0088 -		ISM02.4	PPB	4/23/2019 14:21	SMPL
K2474-03	MA5131D	304	0.0063 -		ISM02.4	PPB	4/23/2019 14:24	SMPL
K2474-04	MA5131S	16051	3.0328 -		ISM02.4	PPB	4/23/2019 14:29	SMPL
CCV084	CCV084	26077	4.9598 -		ISM02.4	PPB	4/23/2019 14:31	SMPL
CCB084	CCB084	175	-0.0185 -		ISM02.4	PPB	4/23/2019 14:34	SMPL