



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102305

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/23/2019	17:30
2	CAL2	1	Water	NA	NA	21.1	7.00	04/23/2019	17:31
3	CAL3	1	Water	NA	NA	21.1	10.03	04/23/2019	17:32
4	ICV	1	Water	NA	NA	21.2	6.99	04/23/2019	17:33
5	CCV1	1	Water	NA	NA	21.1	2.01	04/23/2019	17:35
6	K2200-08	1	Solid	20.03	20	24.8	8.66	04/23/2019	17:36
7	K2200-10	1	Solid	20.04	20	24.6	9.75	04/23/2019	17:37
8	K2200-12	1	Solid	20.02	20	24.7	8.90	04/23/2019	17:39
9	K2541-02	1	Solid	20.03	20	24.1	8.29	04/23/2019	17:40
10	K2542-04	1	Solid	20.02	20	23.7	6.75	04/23/2019	17:41
11	K2542-08	1	Solid	20.03	20	24.0	6.41	04/23/2019	17:42
12	K2542-12	1	Solid	20.04	20	23.4	6.20	04/23/2019	17:44
13	K2542-16	1	Solid	20.01	20	23.8	6.69	04/23/2019	17:45
14	K2542-20	1	Solid	20.04	20	24.2	7.64	04/23/2019	17:46
15	K2542-20DUP	1	Solid	20.03	20	24.1	7.65	04/23/2019	17:48
16	CCV2	1	Water	NA	NA	21.2	12.02	04/23/2019	17:49

13 102305

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2541

WorkList ID : 125205

Date : 04-23-2019 14:17:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/23/2019	Solid	K2200-08	Corrosivity	Cool 4 deg C	PSEG05		EO-02-042319-A	04/23/2019	9045D
04/23/2019	Solid	K2200-10	Corrosivity	Cool 4 deg C	PSEG05		EO-02-042319-B	04/23/2019	9045D
04/23/2019	Solid	K2200-12	Corrosivity	Cool 4 deg C	PSEG05		EO-02-042319-C	04/23/2019	9045D
04/23/2019	Solid	K2541-02	Corrosivity	Cool 4 deg C	PSEG01	K21	CHRT-26835	04/23/2019	9045D
04/23/2019	Solid	K2542-04	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-6	04/23/2019	9045D
04/23/2019	Solid	K2542-08	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-1	04/23/2019	9045D
04/23/2019	Solid	K2542-12	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-7	04/23/2019	9045D
04/23/2019	Solid	K2542-16	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-11	04/23/2019	9045D
04/23/2019	Solid	K2542-20	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-8	04/23/2019	9045D

Date/Time 04-23-19 16:10
Received by: 
Relinquished by:

Date/Time 04-23-19 18:20
Received by: 
Relinquished by: 