



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102316

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP74053
Turbidity std, 40NTU	WP74048
turbidity std 5 NTU	WP74051
turbidity std 20 NTU	WP74049
Turbidity std 1 NTU	WP74052
turbidity std 0 NTU	WP74047
turbidity std 10 NTU	WP74050
10 NTU Standard 500 ml	W2297

Intercept: 0.1025

Slope: 0.965

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.296	04/24/2019	11:25
2	CAL2	1	1	0.996	04/24/2019	11:31
3	CAL3	5	1	4.860	04/24/2019	11:35
4	CAL4	10	1	9.800	04/24/2019	11:41
5	CAL5	20	1	19.200	04/24/2019	11:46
6	CAL6	40	1	38.800	04/24/2019	11:50



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102316

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.970	10.225	04/24/2019	11:51
2	ICB		1	0.301	0.206	04/24/2019	11:56
3	CCV1	10	1	9.780	10.028	04/24/2019	12:07
4	CCB1		1	0.299	0.204	04/24/2019	12:12
5	LB102316BL		1	0.311	0.216	04/24/2019	12:17
6	K2241-07		1	0.839	0.763	04/24/2019	12:23
7	K2241-08		1	2.700	2.692	04/24/2019	12:29
8	CCV2	10	1	9.820	10.070	04/24/2019	12:34
9	CCB2		1	0.292	0.196	04/24/2019	12:39