



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102317

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP74053
Turbidity std, 40NTU	WP74048
turbidity std 5 NTU	WP74051
turbidity std 20 NTU	WP74049
Turbidity std 1 NTU	WP74052
turbidity std 0 NTU	WP74047
turbidity std 10 NTU	WP74050
10 NTU Standard 500 ml	W2297
TURBIDITY LOD STD, 0.5NTU	WP74054
TURBIDITY LOQ STD, 2.0NTU	WP74055

Intercept: 0.1025

Slope: 0.965

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.296	04/24/2019	11:25
2	CAL2	1	1	0.996	04/24/2019	11:31
3	CAL3	5	1	4.860	04/24/2019	11:35
4	CAL4	10	1	9.800	04/24/2019	11:32
5	CAL5	20	1	19.200	04/24/2019	11:46
6	CAL6	40	1	38.800	04/24/2019	11:50



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102317

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.970	10.225	04/24/2019	11:51
2	ICB		1	0.301	0.206	04/24/2019	11:56
3	CCV1	10	1	9.820	10.070	04/24/2019	12:34
4	CCB1		1	0.292	0.196	04/24/2019	12:39
5	LB102317BL		1	0.308	0.213	04/24/2019	12:45
6	K2241-10		1	0.830	0.754	04/24/2019	12:49
7	K2241-11		1	2.790	2.785	04/24/2019	12:55
8	CCV2	10	1	9.760	10.008	04/24/2019	13:05
9	CCB2		1	0.297	0.202	04/24/2019	13:09