


LB102324

7471B

INSTRUMENT ID : C01

Linear



A= 0.0000e+000

B= 1.6571e-004

C= 1.6659e-002

Rho= 0.9999691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	59	0.000	59				
0.2	0.200	0.209	0.009	1160	0.0 %	1160				
2.5	2.500	2.438	-0.062	14613	0.0 %	14613				
5.0	5.000	5.018	0.018	30182	0.0 %	30182				
7.5	7.500	7.511	0.011	45224	0.0 %	45224				
10.0	10.000	9.998	-0.002	60232	0.0 %	60232				

LB102324 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Apr 2019 14:02:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	59	-	0.0000	7471B	PPB	24 Apr 2019 14:09:30	S	Std
0.2 - 1	50-2	1160	-	0.2000	7471B	PPB	24 Apr 2019 14:11:47	S	Std
2.5 - 1	52-5	14613	-	2.5000	7471B	PPB	24 Apr 2019 14:14:04	S	Std
5.0 - 1	55-0	30182	-	5.0000	7471B	PPB	24 Apr 2019 14:16:21	S	Std
7.5 - 1	57-5	45224	-	7.5000	7471B	PPB	24 Apr 2019 14:18:38	S	Std
10.0 - 1	510-0	60232	-	10.0000	7471B	PPB	24 Apr 2019 14:25:01	S	Std
ICV19 - 1	ICV19	24277	4.0396	-	7471B	PPB	24 Apr 2019 14:28:38	U	SMPL
ICB19 - 1	ICB19	-16	0.0140	-	7471B	PPB	24 Apr 2019 14:30:53	U	SMPL
CCV91 - 1	CCV91	30143	5.0116	-	7471B	PPB	24 Apr 2019 14:33:09	U	SMPL
CCB91 - 1	CCB91	-62	0.0064	-	7471B	PPB	24 Apr 2019 14:35:24	U	SMPL
CRA - 1	CRA	1154	0.2079	-	7471B	PPB	24 Apr 2019 14:37:40	U	SMPL
HighStd - 1	HighStd	60759	10.0850	-	7471B	PPB	24 Apr 2019 14:39:55	U	SMPL
ChkStd - 1	ChkStd	42948	7.1336	-	7471B	PPB	24 Apr 2019 14:42:11	U	SMPL
PB119181BL - 1	PBS	-53	0.0079	-	7471B	PPB	24 Apr 2019 14:44:29	U	SMPL
PB119181BS - 1	LCSS	21668	3.6073	-	7471B	PPB	24 Apr 2019 14:46:46	U	SMPL
K2200-07 - 1	EO-02-042319-A	32141	5.3427	-	7471B	PPB	24 Apr 2019 14:49:02	U	SMPL
K2200-09 - 1	EO-02-042319-B	105678	17.5285	-	7471B	PPB	24 Apr 2019 14:51:19	U	SMPL
K2200-11 - 1	EO-02-042319-C	13952	2.3286	-	7471B	PPB	24 Apr 2019 14:53:38	U	SMPL
K2539-01 - 1	FK-SF-03-006-A	10339	1.7299	-	7471B	PPB	24 Apr 2019 14:55:57	U	SMPL
K2539-03 - 1	FK-SF-03-006-B	11682	1.9525	-	7471B	PPB	24 Apr 2019 14:58:13	U	SMPL
CCV92 - 1	CCV92	30317	5.0405	-	7471B	PPB	24 Apr 2019 15:00:28	U	SMPL
CCB92 - 1	CCB92	-55	0.0075	-	7471B	PPB	24 Apr 2019 15:02:44	U	SMPL
K2540-01 - 1	FK-SF-03-007-A	9847	1.6484	-	7471B	PPB	24 Apr 2019 15:04:59	U	SMPL
K2540-03 - 1	FK-SF-03-007-B	9458	1.5839	-	7471B	PPB	24 Apr 2019 15:07:15	U	SMPL
K2540-03DUP - 1	FK-SF-03-007-BDUP	9609	1.6090	-	7471B	PPB	24 Apr 2019 15:09:30	U	SMPL
K2540-03MS - 1	FK-SF-03-007-BMS	32102	5.3363	-	7471B	PPB	24 Apr 2019 15:11:46	U	SMPL
K2540-03MSD - 1	FK-SF-03-007-BMSD	32041	5.3262	-	7471B	PPB	24 Apr 2019 15:14:02	U	SMPL
K2542-03 - 1	TP-6	2215	0.3837	-	7471B	PPB	24 Apr 2019 15:16:19	U	SMPL
K2542-07 - 1	TP-1	1948	0.3395	-	7471B	PPB	24 Apr 2019 15:18:38	U	SMPL
K2542-11 - 1	TP-7	4995	0.8444	-	7471B	PPB	24 Apr 2019 15:20:55	U	SMPL
K2542-15 - 1	TP-11	1925	0.3357	-	7471B	PPB	24 Apr 2019 15:23:12	U	SMPL
K2542-19 - 1	TP-8	5744	0.9685	-	7471B	PPB	24 Apr 2019 15:25:27	U	SMPL
CCV93 - 1	CCV93	30455	5.0633	-	7471B	PPB	24 Apr 2019 15:27:43	U	SMPL
CCB93 - 1	CCB93	-9	0.0152	-	7471B	PPB	24 Apr 2019 15:29:58	U	SMPL
K2540-03LX5 - 1		1877	0.3277	-	7471B	PPB	24 Apr 2019 15:32:15	U	SMPL
K2540-03A - 1		32183	5.3497	-	7471B	PPB	24 Apr 2019 15:34:30	U	SMPL
K2200-09DLX5 - 1		21351	3.5547	-	7471B	PPB	24 Apr 2019 15:36:46	U	SMPL
CCV94 - 1	CCV94	30940	5.1437	-	7471B	PPB	24 Apr 2019 15:39:02	U	SMPL
CCB94 - 1	CCB94	-10	0.0150	-	7471B	PPB	24 Apr 2019 15:41:18	U	SMPL