



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB102327

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/24/2019	12:45
2	K2538-01		1	9	100.00	>212.0	>212.0	04/24/2019	13:20
3	K2538-01DUP		1	9	100.00	>212.0	>212.0	04/24/2019	13:50
4	K2541-01		1	7	100.00	>212.0	>212.0	04/24/2019	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102327

WorkList Name : FP42419

WorkList ID : 125232

Date : 04-24-2019 09:04:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/06/2019	Water	K2538-01	Flash Point	Cool 4 deg C	PSEG01	K11	NB-21926	04/23/2019	1010A
05/06/2019	Water	K2541-01	Flash Point	Cool 4 deg C	PSEG01	K21	DRUM-COMP	04/23/2019	1010A

Date/Time 04/24/19 12:15
 Received by: KS
 Relinquished by: CP

Date/Time 04/24/19 14:25
 Received by: CP
 Relinquished by: KS