



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB102334
Analysis Date: 04/25/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Aparna
Extraction Date: 04/25/2019
Extraction IN Time: 07:25
Extraction OUT Time: 08:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB102334BL	LB102334BL	WATER	1.3	1000	100	2.1540	2.1540	3.02	2.1540	2.1540	0.0000	0
2	LB102334BS	LB102334BS	WATER	1.3	1000	100	2.5698	2.5698	3.01	2.5888	2.5888	0.0190	19
3	LB102334BSD	LB102334BSD	WATER	1.3	1000	100	2.1036	2.1036	3.03	2.1228	2.1228	0.0192	19.2
4	K2545-01	FMI-133	WATER	1.3	1000	100	3.0339	3.0339	3.03	3.0344	3.0344	0.0005	0.5
5	K2546-01	GRAB	WATER	1.3	1000	100	3.1733	3.1733	3.05	3.1793	3.1793	0.0060	6



QC Batch# LB102334 Test: TPH Analysis Date: 04/25/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2515
pH Paper 0-14	M2540
Sodium Sulfate	EP1944
1:1 HCL	WP69312
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP71442
LCSWD	10.00 ML	WP71443
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>09:41</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:40</u>		
Bal Check Time:	<u>07:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>10:15</u>
		Out Time:	<u>09:40</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>11:16</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:40</u>		
Bal Check Time:	<u>12:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>11:45</u>
		Out Time:	<u>11:15</u>		

WORKLIST(Hardcopy Internal Chain)

102332 JP
102334 JP

WorkList Name : TPH K2546

WorkList ID : 125261

Date : 04-25-2019 7:15:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/29/2019	Water	K2545-01-✓	TPH	Conc H2SO4 to	PSEG01	K11	FMI-133	04/24/2019	1664A
05/01/2019	Water	K2546-01 ✓	TPH	Cool 4 deg C	ARAM01	K11	GRAB	04/24/2019	1664A

Date/Time 04.25.19 7:20
Received by: JP
Relinquished by: CN

Date/Time 04.25.19 09:15
Received by: CN
Relinquished by: JP