

LB102340


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44579



A= 0.0000e+000

B= 2.2689e-004

C= -7.0541e-002

Rho= 0.9999275

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	300	0.000	300				
0.2	0.200	0.184	-0.016	1122	0.0 %	1122				
2.5	2.500	2.541	0.041	11511	0.0 %	11511				
5.0	5.000	5.020	0.020	22438	0.0 %	22438				
7.5	7.500	7.413	-0.087	32984	0.0 %	32984				
10.0	10.000	10.044	0.044	44579	0.0 %	44579				

LB102340 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 25 Apr 2019 11:21:05		Date		Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	50-0	300	-	0.0000	7470A	PPB	25 Apr 2019 11:24:24	S	Std
0.2 - 1	50-2	1122	-	0.2000	7470A	PPB	25 Apr 2019 11:26:32	S	Std
2.5 - 1	52-5	11511	-	2.5000	7470A	PPB	25 Apr 2019 11:28:40	S	Std
5.0 - 1	55-0	22438	-	5.0000	7470A	PPB	25 Apr 2019 11:30:48	S	Std
7.5 - 1	57-5	32984	-	7.5000	7470A	PPB	25 Apr 2019 11:32:56	S	Std
10.0 - 1	510-0	44579	-	10.0000	7470A	PPB	25 Apr 2019 11:35:03	S	Std
ICV20 - 1	ICV20	16857	3.7541	-	7470A	PPB	25 Apr 2019 11:39:29	U	SMPL
ICB20 - 1	ICB20	211	-0.0227	-	7470A	PPB	25 Apr 2019 11:41:35	U	SMPL
CCV95 - 1	CCV95	22314	4.9922	-	7470A	PPB	25 Apr 2019 11:43:41	U	SMPL
CCB95 - 1	CCB95	190	-0.0274	-	7470A	PPB	25 Apr 2019 11:45:47	U	SMPL
CRA - 1	CRA	1079	0.1743	-	7470A	PPB	25 Apr 2019 11:47:53	U	SMPL
HighStd - 1	HighStd	43542	9.8086	-	7470A	PPB	25 Apr 2019 11:49:59	U	SMPL
ChkStd - 1	ChkStd	30947	6.9509	-	7470A	PPB	25 Apr 2019 11:52:06	U	SMPL
PB119213BL - 1	PBW	448	0.0311	-	7470A	PPB	25 Apr 2019 11:54:12	U	SMPL
PB119213BS - 1	LCSW	18016	4.0171	-	7470A	PPB	25 Apr 2019 11:56:20	U	SMPL
K2200-08 - 1	EO-02-042319-A	244	-0.0152	-	7470A	PPB	25 Apr 2019 11:58:27	U	SMPL
K2200-08DUP - 1	EO-02-042319-ADUP	269	-0.0095	-	7470A	PPB	25 Apr 2019 12:00:35	U	SMPL
K2200-08MS - 1	EO-02-042319-AMS	17778	3.9631	-	7470A	PPB	25 Apr 2019 12:02:43	U	SMPL
K2200-08MSD - 1	EO-02-042319-AMSD	17929	3.9973	-	7470A	PPB	25 Apr 2019 12:04:52	U	SMPL
K2200-10 - 1	EO-02-042319-B	564	0.0574	-	7470A	PPB	25 Apr 2019 12:06:58	U	SMPL
CCV96 - 1	CCV96	22778	5.0975	-	7470A	PPB	25 Apr 2019 12:09:05	U	SMPL
CCB96 - 1	CCB96	225	-0.0195	-	7470A	PPB	25 Apr 2019 12:11:11	U	SMPL
K2200-12 - 1	EO-02-042319-C	398	0.0198	-	7470A	PPB	25 Apr 2019 12:13:17	U	SMPL
K2541-02 - 1	CHRT-26835	155	-0.0354	-	7470A	PPB	25 Apr 2019 12:15:24	U	SMPL
K2542-04 - 1	TP-6	125	-0.0422	-	7470A	PPB	25 Apr 2019 12:17:30	U	SMPL
K2542-08 - 1	TP-1	281	-0.0068	-	7470A	PPB	25 Apr 2019 12:19:37	U	SMPL
K2542-12 - 1	TP-7	456	0.0329	-	7470A	PPB	25 Apr 2019 12:21:44	U	SMPL
K2542-16 - 1	TP-11	590	0.0633	-	7470A	PPB	25 Apr 2019 12:23:51	U	SMPL
K2542-20 - 1	TP-8	539	0.0518	-	7470A	PPB	25 Apr 2019 12:25:59	U	SMPL
PB119213TB - 1	PB119158TB	339	0.0064	-	7470A	PPB	25 Apr 2019 12:28:06	U	SMPL
K2200-08LX5 - 1		37	-0.0621	-	7470A	PPB	25 Apr 2019 12:30:14	U	SMPL
K2200-08A - 1		17720	3.9499	-	7470A	PPB	25 Apr 2019 12:32:21	U	SMPL
CCV97 - 1	CCV97	23262	5.2073	-	7470A	PPB	25 Apr 2019 12:34:28	U	SMPL
CCB97 - 1	CCB97	257	-0.0122	-	7470A	PPB	25 Apr 2019 12:36:37	U	SMPL