



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102341

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/25/2019	11:10
2	CAL2	1	Water	NA	NA	21.1	7.00	04/25/2019	11:11
3	CAL3	1	Water	NA	NA	21.1	10.03	04/25/2019	11:12
4	ICV	1	Water	NA	NA	21.2	6.99	04/25/2019	11:15
5	CCV1	1	Water	NA	NA	21.1	2.01	04/25/2019	11:16
6	K2567-01	1	Water	NA	NA	21.8	4.71	04/25/2019	11:18
7	K2567-01DUP	1	Water	NA	NA	21.7	4.72	04/25/2019	11:20
8	K2567-02	1	Water	NA	NA	21.7	4.78	04/25/2019	11:22
9	CCV2	1	Water	NA	NA	21.2	12.01	04/25/2019	11:24

13102341

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2567

WorkList ID : 125312

Date : 04-25-2019 11:04:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2019	Water	K2567-01 - 1A	pH	Cool 4 deg C	STER16	K61	MAIN-CELL	04/24/2019	SM 4500-H B
04/24/2019	Water	K2567-02 - 1A	pH	Cool 4 deg C	STER16	K61	EX-CELL	04/24/2019	SM 4500-H B

Date/Time 04.25.19 11:05
Received by: Jg
Relinquished by: cn

Date/Time 04.25.19 12:05
Received by: cn
Relinquished by: Jg