



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Aparna

Run Number: LB102344

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/25/2019	09:50
2	K2538-06		1	8	100.00	>212.0	>212.0	04/25/2019	10:20
3	K2538-06DUP		1	8	100.00	>212.0	>212.0	04/25/2019	10:45
4	K2545-02		1	9	100.00	>212.0	>212.0	04/25/2019	11:15
5	K2550-01		1	9	100.00	>212.0	>212.0	04/25/2019	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102344

WorkList Name : FP42519

WorkList ID : 125290

Date : 04-25-2019 09:15:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/06/2019	Water	K2538-06 B	Flash Point	Cool 4 deg C	PSEG01	K11	NB-20523	04/23/2019	1010A
04/29/2019	Water	K2545-02 C	Flash Point	Cool 4 deg C	PSEG01	K11	FMI-133	04/24/2019	1010A
05/07/2019	Water	K2550-01 L	Flash Point	Cool 4 deg C	DIST01	K11	55-ST-10	04/24/2019	1010A

Date/Time 04/25/19 09:20
 Received by: KS
 Relinquished by: CP

Date/Time 04/25/19 11:50
 Received by: CP
 Relinquished by: KS