



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102357

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/25/2019	16:00
2	CAL2	1	Water	NA	NA	21.1	7.01	04/25/2019	16:01
3	CAL3	1	Water	NA	NA	21.1	10.02	04/25/2019	16:02
4	ICV	1	Water	NA	NA	21.1	6.99	04/25/2019	16:03
5	CCV1	1	Water	NA	NA	21.1	2.02	04/25/2019	16:05
6	K2545-01	1	Water	NA	NA	21.3	9.58	04/25/2019	16:06
7	K2545-01DUP	1	Water	NA	NA	21.4	9.59	04/25/2019	16:07
8	K2578-01	1	Water	NA	NA	20.0	7.60	04/25/2019	16:09
9	CCV2	1	Water	NA	NA	21.1	12.01	04/25/2019	16:10

WORKLIST(Hardcopy Internal Chain)

1B102357

WorkList Name : PH W K2578

WorkList ID : 125321

Date : 04-25-2019 15:48:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2019	Water	K2545-01	pH	Cool 4 deg C	PSEG01	K11	FMI-133	04/24/2019	9040C
04/25/2019	Water	K2578-01	pH	Cool 4 deg C	PSEG01	K11	WATER-COMP	04/25/2019	9040C

Date/Time 04.25.19 15:50
 Received by: JD
 Relinquished by: CD

Date/Time 04.25.19 16:30
 Received by: CD
 Relinquished by: JD