



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102372

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP74171
Turbidity std, 40NTU	WP74165
turbidity std 5 NTU	WP74168
turbidity std 20 NTU	WP74166
Turbidity std 1 NTU	WP74169
turbidity std 0 NTU	WP74170
turbidity std 10 NTU	WP74167

Intercept: 0.0229

Slope: 0.979

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.301	04/26/2019	10:28
2	CAL2	1	1	0.994	04/26/2019	10:32
3	CAL3	5	1	4.860	04/26/2019	10:36
4	CAL4	10	1	9.790	04/26/2019	10:40
5	CAL5	20	1	19.200	04/26/2019	10:46
6	CAL6	40	1	39.400	04/26/2019	10:54



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB102372

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.940	10.130	04/26/2019	10:59
2	ICB		1	0.298	0.281	04/26/2019	11:05
3	CCV1	10	1	9.900	10.089	04/26/2019	11:11
4	CCB1		1	0.302	0.285	04/26/2019	11:17
5	LB102372BL		1	0.296	0.279	04/26/2019	11:22
6	K2241-09		1	0.652	0.643	04/26/2019	11:28
7	CCV2	10	1	9.880	10.069	04/26/2019	11:33
8	CCB2		1	0.298	0.281	04/26/2019	11:37