

LB102374


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

62199



A= 0.0000e+000

B= 1.5983e-004

C= -1.0785e-001

Rho= 0.9995368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.088	-0.088	126	0.000	126				
0.2	0.200	0.185	-0.015	1833	0.0 %	1833				
2.5	2.500	2.583	0.083	16838	0.0 %	16838				
5.0	5.000	5.004	0.004	31980	0.0 %	31980				
7.5	7.500	7.682	0.182	48737	0.0 %	48737				
10.0	10.000	9.834	-0.166	62199	0.0 %	62199				

LB102374 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Apr 2019 11:57:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	126	-	0.0000	7471B	PPB	26 Apr 2019 11:59:00	S	Std
0.2 - 1	50-2	1833	-	0.2000	7471B	PPB	26 Apr 2019 12:01:17	S	Std
2.5 - 1	52-5	16838	-	2.5000	7471B	PPB	26 Apr 2019 12:03:34	S	Std
5.0 - 1	55-0	31980	-	5.0000	7471B	PPB	26 Apr 2019 12:05:51	S	Std
7.5 - 1	57-5	48737	-	7.5000	7471B	PPB	26 Apr 2019 12:08:08	S	Std
10.0 - 1	510-0	62199	-	10.0000	7471B	PPB	26 Apr 2019 12:13:32	S	Std
ICV21 - 1	ICV21	25236	3.9257	-	7471B	PPB	26 Apr 2019 12:17:09	U	SMPL
ICB21 - 1	ICB21	24	-0.1040	-	7471B	PPB	26 Apr 2019 12:22:10	U	SMPL
CCV98 - 1	CCV98	31752	4.9672	-	7471B	PPB	26 Apr 2019 12:24:26	U	SMPL
CCB98 - 1	CCB98	-18	-0.1107	-	7471B	PPB	26 Apr 2019 12:26:40	U	SMPL
CRA - 1	CRA	1819	0.1829	-	7471B	PPB	26 Apr 2019 12:28:56	U	SMPL
HighStd - 1	HighStd	62725	9.9177	-	7471B	PPB	26 Apr 2019 12:31:11	U	SMPL
ChkStd - 1	ChkStd	45417	7.1513	-	7471B	PPB	26 Apr 2019 12:33:27	U	SMPL
PB119250BL - 1	PBS	86	-0.0941	-	7471B	PPB	26 Apr 2019 12:35:46	U	SMPL
PB119250BS - 1	LCSS	24042	3.7349	-	7471B	PPB	26 Apr 2019 12:38:02	U	SMPL
K2198-07 - 1	OK-01-042519	3060	0.3812	-	7471B	PPB	26 Apr 2019 12:40:18	U	SMPL
K2203-17 - 1	SU-03-042419-A	1364	0.1102	-	7471B	PPB	26 Apr 2019 12:42:35	U	SMPL
K2203-17DUP - 1	SU-03-042419-ADUP	1416	0.1185	-	7471B	PPB	26 Apr 2019 12:44:52	U	SMPL
K2203-17MS - 1	SU-03-042419-AMS	26421	4.1151	-	7471B	PPB	26 Apr 2019 12:47:10	U	SMPL
K2203-17MSD - 1	SU-03-042419-AMSD	26371	4.1071	-	7471B	PPB	26 Apr 2019 12:49:25	U	SMPL
CCV99 - 1	CCV99	32333	5.0600	-	7471B	PPB	26 Apr 2019 12:51:41	U	SMPL
CCB99 - 1	CCB99	25	-0.1039	-	7471B	PPB	26 Apr 2019 12:53:57	U	SMPL
K2203-19 - 1	SU-03-042419-B	2320	0.2630	-	7471B	PPB	26 Apr 2019 12:56:13	U	SMPL
K2544-01 - 1	135-20	47596	7.4996	-	7471B	PPB	26 Apr 2019 12:58:29	U	SMPL
K2553-01 - 1	RBR200034	2592	0.3064	-	7471B	PPB	26 Apr 2019 13:00:46	U	SMPL
K2553-04 - 1	RBR200051	1667	0.1586	-	7471B	PPB	26 Apr 2019 13:03:02	U	SMPL
K2554-01 - 1	RT3860	1851	0.1880	-	7471B	PPB	26 Apr 2019 13:05:18	U	SMPL
K2564-01 - 1	C-1_042419	37868	5.9447	-	7471B	PPB	26 Apr 2019 13:07:35	U	SMPL
K2564-02 - 1	C-2_042419	60839	9.6162	-	7471B	PPB	26 Apr 2019 13:09:51	U	SMPL
K2564-03 - 1	C-3_042419	81413	12.9046	-	7471B	PPB	26 Apr 2019 13:12:10	U	SMPL
K2571-01 - 1	144-PENNINGTON	4902	0.6757	-	7471B	PPB	26 Apr 2019 13:14:28	U	SMPL
K2575-01 - 1	PAINT-MATERIAL-DRUM	5055	0.7001	-	7471B	PPB	26 Apr 2019 13:16:44	U	SMPL
CCV01 - 1	CCV01	31924	4.9947	-	7471B	PPB	26 Apr 2019 13:19:00	U	SMPL
CCB01 - 1	CCB01	-34	-0.1133	-	7471B	PPB	26 Apr 2019 13:21:16	U	SMPL
K2577-01 - 1	20190303	1679	0.1605	-	7471B	PPB	26 Apr 2019 13:23:32	U	SMPL
K2577-02 - 1	20190304	2246	0.2511	-	7471B	PPB	26 Apr 2019 13:25:48	U	SMPL
K2582-01 - 1	SSSI-0419-S-DH-6-T	54583	8.6163	-	7471B	PPB	26 Apr 2019 13:28:03	U	SMPL
K2582-03 - 1	SSSI-0419-S-DH-6-B	5992	0.8499	-	7471B	PPB	26 Apr 2019 13:30:19	U	SMPL
K2203-17LX5 - 1		399	-0.0441	-	7471B	PPB	26 Apr 2019 13:32:35	U	SMPL
K2203-17A - 1		26621	4.1471	-	7471B	PPB	26 Apr 2019 13:34:51	U	SMPL
K2564-03DLX4 - 1		19955	3.0816	-	7471B	PPB	26 Apr 2019 13:37:08	U	SMPL
CCV02 - 1	CCV02	32091	5.0214	-	7471B	PPB	26 Apr 2019 13:39:25	U	SMPL
CCB02 - 1	CCB02	146	-0.0845	-	7471B	PPB	26 Apr 2019 13:41:42	U	SMPL